

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071222\
 Data File : PQ058554.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jul 2022 00:06
 Operator : YP\AJ
 Sample : AR1254ICC1600
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12545005

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 13 06:13:18 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ071222CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 13 06:13:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.688	4.050	688.7E6	561.5E6	72.965	79.703
2)	SA Decachlor...	10.598	9.174	1017.3E6	580.8E6	138.180	148.908
Target Compounds							
26)	L6 AR-1254-1	6.726	5.957	293.7E6	464.8E6	1410.733	1537.237
27)	L6 AR-1254-2	6.944	6.102	445.5E6	393.1E6	1431.155	1512.212
28)	L6 AR-1254-3	7.314	6.512	519.4E6	625.5E6	1448.753	1561.606
29)	L6 AR-1254-4	7.600	6.738	312.5E6	396.4E6	1439.808	1547.995
30)	L6 AR-1254-5	8.020	7.158	365.2E6	478.1E6	1403.036	1567.324

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071222\
Data File : PQ058554.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jul 2022 00:06
Operator : YP\AJ
Sample : AR1254ICC1600
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12545005

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 13 06:13:18 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ071222CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 13 06:13:10 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

