

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071318\
 Data File : PQ030669.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jul 2018 00:25
 Operator : SM\SJ
 Sample : AR1242 LOD
 Misc : 25PPB LOD WATER ECD_Q
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1242 LOD

Manual Integrations
 APPROVED

Sohil
 7/18/2018 4:54:31 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 05:27:01 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ071218.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 12 12:41:01 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.425	3.691	788.4E6	234.4E6	17.556	15.070
2) SA Decachlor...	10.094	8.652	627.0E6	363.6E6	19.186	17.371
Target Compounds						
16) L4 AR-1242-1	5.323	4.548	22106699	6514569	27.756m	20.224m#
17) L4 AR-1242-2	5.587	4.767	42105215	10383340	32.616m	21.304m#
18) L4 AR-1242-3	5.673	4.956	27567646	6635761	26.891	22.270m
19) L4 AR-1242-4	6.063	5.039	17328310	6045544	18.848m	19.746m
20) L4 AR-1242-5	6.195	5.209	25724545	6107740	25.809m	17.536m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071318\
Data File : PQ030669.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jul 2018 00:25
Operator : SM\SJ
Sample : AR1242 LOD
Misc : 25PPB LOD WATER ECD_Q
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
AR1242 LOD

Manual Integrations
APPROVED

Sohil
7/18/2018 4:54:31 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 05:27:01 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ071218.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 12 12:41:01 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

