

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071420\
 Data File : PQ048683.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jul 2020 17:21
 Operator : DD\AJ
 Sample : L3287-03
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 FK-GF-03-012-B

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 14 18:19:28 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ070820.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 09 01:18:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.277	4.289	244.2E6	128.4E6	35.069	36.846
2)	SA Decachlor...	11.534	9.652	206.3E6	95980829	26.867	23.852

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071420\
Data File : PQ048683.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jul 2020 17:21
Operator : DD\AJ
Sample : L3287-03
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
FK-GF-03-012-B

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 14 18:19:28 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ070820.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 09 01:18:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.277 min
Delta R.T.: -0.001 min
Response: 244218406
Conc: 35.07 ng/ml

Instrument :
ECD_Q
ClientSampleId :
FK-GF-03-012-B




#1 Tetrachloro-m-xylene

R.T.: 4.289 min
Delta R.T.: 0.004 min
Response: 128368127
Conc: 36.85 ng/ml



#2 Decachlorobiphenyl

R.T.: 11.534 min
Delta R.T.: -0.004 min
Response: 206286829
Conc: 26.87 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.652 min
Delta R.T.: 0.000 min
Response: 95980829
Conc: 23.85 ng/ml