

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071519\
 Data File : PQ041206.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jul 2019 13:35
 Operator : SM\AJ
 Sample : K3782-08
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LC-TB03-190711

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 15 14:45:40 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ071219.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jul 13 02:14:45 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.716	4.771	211.3E6	137.4E6	20.806	21.078
2)	SA Decachlor...	12.134	10.544	182.3E6	67889458	14.197	14.522

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071519\
Data File : PQ041206.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jul 2019 13:35
Operator : SM\AJ
Sample : K3782-08
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LC-TB03-190711

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 15 14:45:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ071219.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jul 13 02:14:45 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.716 min
Delta R.T.: -0.001 min
Response: 211347133
Conc: 20.81 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LC-TB03-190711




#1 Tetrachloro-m-xylene

R.T.: 4.771 min
Delta R.T.: 0.000 min
Response: 137447633
Conc: 21.08 ng/ml



#2 Decachlorobiphenyl

R.T.: 12.134 min
Delta R.T.: -0.003 min
Response: 182328261
Conc: 14.20 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.544 min
Delta R.T.: -0.002 min
Response: 67889458
Conc: 14.52 ng/ml