

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071620\
 Data File : PQ048756.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jul 2020 21:47
 Operator : DD\AJ
 Sample : L3216-08
 Misc : AR1262 LOQ 50 PPB
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOQ-WATER-02-QT3-2020

Manual Integrations
 APPROVED

Ankita
 7/17/2020 12:08:01 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 17 01:39:31 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ071620.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 16 05:42:31 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.269	4.283	211.9E6	101.1E6	25.078	24.100
2) SA Decachlor...	11.520	9.646	261.4E6	132.0E6	25.388	24.491
Target Compounds						
36) L8 AR-1262-1	8.928	7.670	32632962	10298556	56.446	51.907
37) L8 AR-1262-2	9.526	8.221	58933129	35027803	53.685	48.031
38) L8 AR-1262-3	9.863	8.509	26960741	16702105	54.861	61.310
39) L8 AR-1262-4	9.948f	8.573	32686117	27926267	55.975m	53.818
40) L8 AR-1262-5	10.695	9.078	24276379	12797677	57.867	55.011

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071620\
Data File : PQ048756.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jul 2020 21:47
Operator : DD\AJ
Sample : L3216-08
Misc : AR1262 LOQ 50 PPB
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2020

Manual Integrations
APPROVED

Ankita
7/17/2020 12:08:01 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 17 01:39:31 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ071620.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 16 05:42:31 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

