

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071923\
 Data File : PQ062379.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jul 2023 09:09
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/20/2023
 Supervised By :Ankita Jodhani 07/20/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 19 16:16:11 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ070123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 04 05:32:27 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.402	2.754	248.4E6	136.2E6	50.928	54.264m
2) SA Decachlor...	8.589	7.526	193.4E6	184.5E6	53.239	55.357
Target Compounds						
3) L1 AR-1016-1	4.504	3.755	86138364	51654509	522.945	521.892
4) L1 AR-1016-2	4.523	3.770	127.2E6	75848678	514.473	526.000
5) L1 AR-1016-3	4.581	3.931	79242104	38470027	519.942	509.924
6) L1 AR-1016-4	4.672	3.980	65751725	34129072	524.565	519.633
7) L1 AR-1016-5	4.959	4.175	63063123	42401851	518.113	517.212
31) L7 AR-1260-1	6.063	5.168	121.9E6	87673183	516.525	526.397
32) L7 AR-1260-2	6.323	5.358	148.6E6	108.0E6	514.500	519.458
33) L7 AR-1260-3	6.677	5.499	92260307	101.9E6	522.495	523.997
34) L7 AR-1260-4	6.895	5.961	111.0E6	77900758	526.874	534.073
35) L7 AR-1260-5	7.207	6.208	219.0E6	179.1E6	533.354	539.226

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ071923\
Data File : PQ062379.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jul 2023 09:09
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/20/2023
Supervised By :Ankita Jodhani 07/20/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 19 16:16:11 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ070123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 04 05:32:27 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

