

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072022\
 Data File : PQ058650.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jul 2022 19:43
 Operator : YP\AJ
 Sample : N3739-02MS
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 H0B39MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 00:08:59 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ071222CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jul 14 12:27:52 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.689	4.053	162.8E6	129.0E6	23.137	23.077
2) SA Decachlor...	10.600	9.176	282.3E6	151.9E6	50.466	45.877
Target Compounds						
3) L1 AR-1016-1	5.869	5.150	74130131	79199571	490.344	507.354
4) L1 AR-1016-2	5.892	5.170	130.1E6	114.4E6	499.899	509.740
5) L1 AR-1016-3	5.954	5.350	84434731	56290622	522.891	509.450
6) L1 AR-1016-4	6.055	5.387	72156994	45332207	535.927	476.329
7) L1 AR-1016-5	6.348	5.605	58528073	57459290	498.875	491.795
31) L7 AR-1260-1	7.474	6.641	91937701	108.6E6	522.081	542.647
32) L7 AR-1260-2	7.731	6.825	106.3E6	124.8E6	507.672	536.185
33) L7 AR-1260-3	8.018	6.984	147.0E6	111.5E6	555.697	521.332
34) L7 AR-1260-4	8.329	7.456	94509218	73920036	464.578	418.244
35) L7 AR-1260-5	8.670	7.693	211.4E6	181.4E6	479.551	456.021

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072022\
Data File : PQ058650.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jul 2022 19:43
Operator : YP\AJ
Sample : N3739-02MS
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
H0B39MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 00:08:59 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ071222CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jul 14 12:27:52 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

