

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ072118.M
 Title : GC EXTRACTABLES
 Last Update : Mon Jul 23 03:18:02 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ030762.D 1000 =PQ030760.D 750 =PQ030761.D
 250 =PQ030763.D 50 =PQ030764.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.808	5.333	6.435	5.207	5.588	5.474	E7	11.08
2) SA	Decachlorobiphenyl	3.349	3.350	3.724	3.915	4.149	3.697	E7	9.51
3) L1	AR-1016-1	7.099	7.847	9.323	7.952	6.686	7.781	E5	12.97
4) L1	AR-1016-2	0.993	1.171	1.331	1.093	1.079	1.134	E6	11.24
5) L1	AR-1016-3	1.268	1.538	1.692	1.353	1.349	1.440	E6	11.96
6) L1	AR-1016-4	1.132	1.357	1.482	1.174	1.243	1.277	E6	11.14
7) L1	AR-1016-5	1.066	1.225	1.427	1.166	1.163	1.210	E6	11.10
8) L2	AR-1221-1	8.185					8.185	E5	0.00
9) L2	AR-1221-2	6.943					6.943	E5	0.00
10) L2	AR-1221-3	1.733					1.733	E6	0.00
11) L3	AR-1232-1	1.206					1.206	E6	0.00
12) L3	AR-1232-2	5.722					5.722	E5	0.00
13) L3	AR-1232-3	7.073					7.073	E5	0.00
14) L3	AR-1232-4	6.299					6.299	E5	0.00
15) L3	AR-1232-5	4.744					4.744	E5	0.00
16) L4	AR-1242-1	9.821					9.821	E5	0.00
17) L4	AR-1242-2	1.507					1.507	E6	0.00
18) L4	AR-1242-3	1.253					1.253	E6	0.00
19) L4	AR-1242-4	1.121					1.121	E6	0.00
20) L4	AR-1242-5	1.339					1.339	E6	0.00
21) L5	AR-1248-1	1.715					1.715	E6	0.00
22) L5	AR-1248-2	1.543					1.543	E6	0.00
23) L5	AR-1248-3	1.454					1.454	E6	0.00
24) L5	AR-1248-4	1.846					1.846	E6	0.00
25) L5	AR-1248-5	1.853					1.853	E6	0.00
26) L6	AR-1254-1	2.284					2.284	E6	0.00
27) L6	AR-1254-2	3.273					3.273	E6	0.00
28) L6	AR-1254-3	3.670					3.670	E6	0.00
29) L6	AR-1254-4	2.620					2.620	E6	0.00
30) L6	AR-1254-5	2.600					2.600	E6	0.00
31) L7	AR-1260-1	1.938	2.094	2.314	2.163	2.268	2.155	E6	6.92
32) L7	AR-1260-2	2.265	2.395	2.605	2.539	2.588	2.478	E6	5.85
33) L7	AR-1260-3	2.527	2.697	2.919	2.917	2.480	2.708	E6	7.67
34) L7	AR-1260-4	1.687	1.788	1.935	1.999	1.726	1.827	E6	7.36
35) L7	AR-1260-5	3.421	3.570	3.918	4.073	3.917	3.780	E6	7.21
36) L8	AR-1262-1	1.888					1.888	E6	0.00
37) L8	AR-1262-2	2.219					2.219	E6	0.00
38) L8	AR-1262-3	2.758					2.758	E6	0.00
39) L8	AR-1262-4	2.397					2.397	E6	0.00
40) L8	AR-1262-5	4.513					4.513	E6	0.00
41) L9	AR-1268-1	4.366					4.366	E6	0.00
42) L9	AR-1268-2	3.932					3.932	E6	0.00
43) L9	AR-1268-3	3.317					3.317	E6	0.00
44) L9	AR-1268-4	1.509					1.509	E6	0.00
45) L9	AR-1268-5	1.130					1.130	E7	0.00

Signal #2 Calibration Files

500 =PQ030762.D 1000 =PQ030760.D 750 =PQ030761.D
 250 =PQ030763.D 50 =PQ030764.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.939	1.881	2.058	2.040	2.059	1.995	E7	4.06
2) SA	Decachlorobiphenyl	2.422	2.349	2.506	2.578	2.713	2.514	E7	5.61
3) L1	AR-1016-1	4.505	4.319	4.624	4.697	4.522	4.534	E5	3.16
4) L1	AR-1016-2	5.092	4.778	5.283	5.488	5.395	5.207	E5	5.41
5) L1	AR-1016-3	3.768	3.593	3.991	4.269	3.547	3.834	E5	7.81

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ072118.M
 Title : GC EXTRACTABLES
 Last Update : Mon Jul 23 03:18:02 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ030762.D 1000 =PQ030760.D 750 =PQ030761.D
 250 =PQ030763.D 50 =PQ030764.D

Compound			500	1000	750	250	50	Avg		%RSD
6)	L1	AR-1016-4	4.739	4.550	4.984	5.206	5.228	4.941	E5	5.98
7)	L1	AR-1016-5	5.465	4.923	5.438	5.802	5.485	5.422	E5	5.83
8)	L2	AR-1221-1	2.571					2.571	E5	0.00
9)	L2	AR-1221-2	2.043					2.043	E5	0.00
10)	L2	AR-1221-3	6.330					6.330	E5	0.00
11)	L3	AR-1232-1	3.684					3.684	E5	0.00
12)	L3	AR-1232-2	2.854					2.854	E5	0.00
13)	L3	AR-1232-3	2.391					2.391	E5	0.00
14)	L3	AR-1232-4	2.297					2.297	E5	0.00
15)	L3	AR-1232-5	2.970					2.970	E5	0.00
16)	L4	AR-1242-1	4.167					4.167	E5	0.00
17)	L4	AR-1242-2	6.406					6.406	E5	0.00
18)	L4	AR-1242-3	4.762					4.762	E5	0.00
19)	L4	AR-1242-4	4.478					4.478	E5	0.00
20)	L4	AR-1242-5	4.850					4.850	E5	0.00
21)	L5	AR-1248-1	6.296					6.296	E5	0.00
22)	L5	AR-1248-2	6.538					6.538	E5	0.00
23)	L5	AR-1248-3	8.131					8.131	E5	0.00
24)	L5	AR-1248-4	1.210					1.210	E6	0.00
25)	L5	AR-1248-5	8.364					8.364	E5	0.00
26)	L6	AR-1254-1	1.301					1.301	E6	0.00
27)	L6	AR-1254-2	1.098					1.098	E6	0.00
28)	L6	AR-1254-3	1.858					1.858	E6	0.00
29)	L6	AR-1254-4	1.243					1.243	E6	0.00
30)	L6	AR-1254-5	7.864					7.864	E5	0.00
31)	L7	AR-1260-1	1.152	1.091	1.169	1.177	1.122	1.142	E6	3.14
32)	L7	AR-1260-2	1.319	1.283	1.363	1.393	1.363	1.344	E6	3.22
33)	L7	AR-1260-3	1.077	1.055	1.128	1.145	1.121	1.105	E6	3.42
34)	L7	AR-1260-4	2.743	2.673	2.894	2.823	2.750	2.777	E6	3.05
35)	L7	AR-1260-5	1.921	1.872	2.042	2.029	2.034	1.980	E6	3.94
36)	L8	AR-1262-1	1.010					1.010	E6	0.00
37)	L8	AR-1262-2	1.194					1.194	E6	0.00
38)	L8	AR-1262-3	1.709					1.709	E6	0.00
39)	L8	AR-1262-4	1.512					1.512	E6	0.00
40)	L8	AR-1262-5	3.264					3.264	E6	0.00
41)	L9	AR-1268-1	3.030					3.030	E6	0.00
42)	L9	AR-1268-2	2.990					2.990	E6	0.00
43)	L9	AR-1268-3	2.538					2.538	E6	0.00
44)	L9	AR-1268-4	1.059					1.059	E6	0.00
45)	L9	AR-1268-5	8.012					8.012	E6	0.00

(#) = Out of Range