

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072123\
 Data File : PQ062469.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Jul 2023 08:55
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/24/2023
 Supervised By :Ankita Jodhani 07/24/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 21 12:40:10 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ070123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 04 05:32:27 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.402	2.753	249.8E6	137.1E6	51.202	54.654m
2) SA Decachlor...	8.590	7.525	194.6E6	178.6E6	53.561	53.585m
Target Compounds						
3) L1 AR-1016-1	4.504	3.755	85728420	49508330	520.456	500.208
4) L1 AR-1016-2	4.523	3.770	127.8E6	72751575	516.730	504.522
5) L1 AR-1016-3	4.580	3.930	79253522	36451307	520.017	483.166
6) L1 AR-1016-4	4.672	3.979	65644619	32891216	523.710	500.786
7) L1 AR-1016-5	4.960	4.174	64364780	40516067	528.807	494.209
31) L7 AR-1260-1	6.063	5.167	122.8E6	82262985	520.571	493.914
32) L7 AR-1260-2	6.324	5.357	149.3E6	103.3E6	516.840	496.778
33) L7 AR-1260-3	6.678	5.498	92290764	98419997	522.668	505.863
34) L7 AR-1260-4	6.896	5.960	111.0E6	75089968	526.965	514.803
35) L7 AR-1260-5	7.208	6.207	220.0E6	174.0E6	535.863	524.008

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072123\
Data File : PQ062469.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Jul 2023 08:55
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/24/2023
Supervised By :Ankita Jodhani 07/24/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 21 12:40:10 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ070123.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 04 05:32:27 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

