

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072225\
 Data File : PQ070577.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2025 10:52
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ072225AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 11:47:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 23 09:47:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.407	2.779	404.0E6	467.9E6	45.991	46.885
2) SA Decachlor...	8.504	7.540	222.1E6	394.1E6	45.288	45.111
Target Compounds						
21) L5 AR-1248-1	4.487	3.780	81175333	97843910	481.623	473.688
22) L5 AR-1248-2	4.751	4.006	112.5E6	147.5E6	487.593	476.627
23) L5 AR-1248-3	4.938	4.043	133.2E6	141.5E6	488.863	474.339
24) L5 AR-1248-4	5.326	4.200	138.0E6	172.6E6	479.564	479.917
25) L5 AR-1248-5	5.359	4.575	139.9E6	163.5E6	495.070	472.090

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072225\
Data File : PQ070577.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2025 10:52
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ072225AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 11:47:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 23 09:47:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

