

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072324\
 Data File : PQ067399.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Jul 2024 14:48
 Operator : YP\AJ
 Sample : AR1232ICC200
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12322020

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 23 15:27:39 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072324CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 23 15:26:51 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.461	2.830	65623734	67891251	10.240	10.110
2)	SA Decachlor...	8.737	7.614	42403820	68052975	20.236	20.326
Target Compounds							
11)	L3 AR-1232-1	3.814	3.176	25708672	28895907	201.814	201.176
12)	L3 AR-1232-2	4.312	3.853	14206871	25834726	201.784	204.558
13)	L3 AR-1232-3	4.594	4.015	25540299	14706906	204.320	200.850
14)	L3 AR-1232-4	4.745	4.101	12260998	14172412	207.321	195.801
15)	L3 AR-1232-5	4.842	4.259	9039237	15345406	197.178	205.657

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072324\
Data File : PQ067399.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Jul 2024 14:48
Operator : YP\AJ
Sample : AR1232ICC200
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12322020

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 23 15:27:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072324CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 23 15:26:51 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

