

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072424\
 Data File : PQ067490.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Jul 2024 17:02
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/25/2024
 Supervised By :Ankita Jodhani 07/25/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 24 17:19:44 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 24 17:19:27 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.460	2.828	300.9E6	324.5E6	50.000	50.000
2) SA Decachlor...	8.733	7.613	200.4E6	292.9E6	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	5.406	4.597	110.7E6	168.5E6	500.000	500.000
27) L6 AR-1254-2	5.627	4.745	168.4E6	152.1E6	500.000	500.000
28) L6 AR-1254-3	5.987	5.128	172.7E6	232.3E6	500.000	500.000
29) L6 AR-1254-4	6.276	5.354	114.2E6	131.4E6	500.000	500.000
30) L6 AR-1254-5	6.695	5.760	144.9E6	214.6E6	508.521m	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072424\
Data File : PQ067490.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Jul 2024 17:02
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/25/2024
Supervised By :Ankita Jodhani 07/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 24 17:19:44 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 24 17:19:27 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

