

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072623\
 Data File : PQ062688.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jul 2023 19:36
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 26 22:30:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 26 22:29:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.403	2.756	357.1E6	241.9E6	74.596	74.388
2)	SA Decachlor...	8.582	7.523	288.2E6	246.1E6	73.846	74.481
Target Compounds							
26)	L6 AR-1254-1	5.326	4.511	138.8E6	116.8E6	738.775	739.258
27)	L6 AR-1254-2	5.543	4.658	219.8E6	101.5E6	739.175	740.274
28)	L6 AR-1254-3	5.896	5.038	236.5E6	161.7E6	739.559	739.768
29)	L6 AR-1254-4	6.181	5.266	177.1E6	106.5E6	740.073	743.658
30)	L6 AR-1254-5	6.594	5.671	188.9E6	151.6E6	737.703	741.832

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072623\
Data File : PQ062688.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jul 2023 19:36
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 26 22:30:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 26 22:29:58 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

