

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072623\
 Data File : PQ062688.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 Jul 2023 19:36
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 26 22:30:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 26 22:29:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.403	2.756	357.1E6	241.9E6	74.596	74.388
2)	SA Decachlor...	8.582	7.523	288.2E6	246.1E6	73.846	74.481
Target Compounds							
26)	L6 AR-1254-1	5.326	4.511	138.8E6	116.8E6	738.775	739.258
27)	L6 AR-1254-2	5.543	4.658	219.8E6	101.5E6	739.175	740.274
28)	L6 AR-1254-3	5.896	5.038	236.5E6	161.7E6	739.559	739.768
29)	L6 AR-1254-4	6.181	5.266	177.1E6	106.5E6	740.073	743.658
30)	L6 AR-1254-5	6.594	5.671	188.9E6	151.6E6	737.703	741.832



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072623\
Data File : PQ062688.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 Jul 2023 19:36
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 26 22:30:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072623.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 26 22:29:58 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.403 min
Delta R.T.: 0.000 min
Response: 357122458
Conc: 74.60 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750




#1 Tetrachloro-m-xylene

R.T.: 2.756 min
Delta R.T.: 0.000 min
Response: 241947716
Conc: 74.39 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.582 min
Delta R.T.: 0.000 min
Response: 288174832
Conc: 73.85 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.523 min
Delta R.T.: 0.000 min
Response: 246079500
Conc: 74.48 ng/ml



#26 AR-1254-1


R.T.: 5.326 min
Delta R.T.: 0.000 min
Response: 138835192
Conc: 738.77 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750




#26 AR-1254-1

R.T.: 4.511 min
Delta R.T.: 0.000 min
Response: 116759220
Conc: 739.26 ng/ml




#27 AR-1254-2

R.T.: 5.543 min
Delta R.T.: 0.000 min
Response: 219766948
Conc: 739.18 ng/ml



#27 AR-1254-2


R.T.: 4.658 min
Delta R.T.: 0.000 min
Response: 101489441
Conc: 740.27 ng/ml



#28 AR-1254-3


R.T.: 5.896 min
Delta R.T.: 0.000 min
Response: 236463459
Conc: 739.56 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750




#28 AR-1254-3

R.T.: 5.038 min
Delta R.T.: 0.000 min
Response: 161709905
Conc: 739.77 ng/ml




#29 AR-1254-4

R.T.: 6.181 min
Delta R.T.: 0.000 min
Response: 177056095
Conc: 740.07 ng/ml



#29 AR-1254-4


R.T.: 5.266 min
Delta R.T.: 0.000 min
Response: 106479838
Conc: 743.66 ng/ml



#30 AR-1254-5

R.T.: 6.594 min
Delta R.T.: 0.000 min
Response: 188874952
Conc: 737.70 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC750



#30 AR-1254-5

R.T.: 5.671 min
Delta R.T.: 0.000 min
Response: 151615146
Conc: 741.83 ng/ml