

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ072718.M
 Title : GC EXTRACTABLES
 Last Update : Mon Jul 30 09:28:37 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ030901.D 1000 =PQ030899.D 750 =PQ030900.D
 250 =PQ030902.D 50 =PQ030903.D

Compound		500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.213	1.233	1.203	1.249	1.368	1.253 E8	5.33
2) SA	Decachlorobiphenyl	2.429	2.464	2.338	2.431	2.686	2.469 E7	5.26
3) L1	AR-1016-1	1.830	1.653	1.577	1.449	1.624	1.627 E6	8.48
4) L1	AR-1016-2	2.260	2.267	2.147	2.475	2.299	2.290 E6	5.17
5) L1	AR-1016-3	4.032	3.983	3.962	4.158	4.328	4.092 E6	3.71
6) L1	AR-1016-4	2.636	2.368	2.367	2.306	2.211	2.377 E6	6.66
7) L1	AR-1016-5	1.805	1.722	1.725	1.720	1.978	1.790 E6	6.20
8) L2	AR-1221-1	1.725					1.725 E6	0.00
9) L2	AR-1221-2	1.217					1.217 E6	0.00
10) L2	AR-1221-3	4.469					4.469 E6	0.00
11) L3	AR-1232-1	2.122					2.122 E6	0.00
12) L3	AR-1232-2	9.436					9.436 E5	0.00
13) L3	AR-1232-3	1.887					1.887 E6	0.00
14) L3	AR-1232-4	1.216					1.216 E6	0.00
15) L3	AR-1232-5	8.020					8.020 E5	0.00
16) L4	AR-1242-1	1.978					1.978 E6	0.00
17) L4	AR-1242-2	2.560					2.560 E6	0.00
18) L4	AR-1242-3	3.944					3.944 E6	0.00
19) L4	AR-1242-4	1.637					1.637 E6	0.00
20) L4	AR-1242-5	1.648					1.648 E6	0.00
21) L5	AR-1248-1	2.779					2.779 E6	0.00
22) L5	AR-1248-2	2.350					2.350 E6	0.00
23) L5	AR-1248-3	2.680					2.680 E6	0.00
24) L5	AR-1248-4	2.611					2.611 E6	0.00
25) L5	AR-1248-5	2.484					2.484 E6	0.00
26) L6	AR-1254-1	2.649					2.649 E6	0.00
27) L6	AR-1254-2	3.770					3.770 E6	0.00
28) L6	AR-1254-3	2.101					2.101 E6	0.00
29) L6	AR-1254-4	2.183					2.183 E6	0.00
30) L6	AR-1254-5	2.467					2.467 E6	0.00
31) L7	AR-1260-1	2.242	2.207	2.099	2.497	2.428	2.295 E6	7.14
32) L7	AR-1260-2	2.358	2.339	2.245	2.387	2.812	2.428 E6	9.10
33) L7	AR-1260-3	2.540	2.581	2.461	2.350	2.134	2.413 E6	7.44
34) L7	AR-1260-4	3.084	3.151	2.956	3.106	4.011	3.261 E6	13.03
35) L7	AR-1260-5	1.831	1.876	1.759	1.915	1.773	1.831 E6	3.61
36) L8	AR-1262-1	1.971					1.971 E6	0.00
37) L8	AR-1262-2	2.067					2.067 E6	0.00
38) L8	AR-1262-3	1.655					1.655 E6	0.00
39) L8	AR-1262-4	2.185					2.185 E6	0.00
40) L8	AR-1262-5	3.820					3.820 E6	0.00
41) L9	AR-1268-1	4.216					4.216 E6	0.00
42) L9	AR-1268-2	3.887					3.887 E6	0.00
43) L9	AR-1268-3	3.151					3.151 E6	0.00
44) L9	AR-1268-4	1.315					1.315 E6	0.00
45) L9	AR-1268-5	9.641					9.641 E6	0.00

Signal #2 Calibration Files

500 =PQ030901.D 1000 =PQ030899.D 750 =PQ030900.D
 250 =PQ030902.D 50 =PQ030903.D

Compound		500	1000	750	250	50	Avg	%RSD
1) SA	Tetrachloro-m-xylene	1.519	1.522	1.512	1.539	1.578	1.534 E7	1.73
2) SA	Decachlorobiphenyl	1.399	1.388	1.347	1.393	1.487	1.403 E7	3.66
3) L1	AR-1016-1	3.587	3.147	3.429	3.155	3.210	3.305 E5	5.88
4) L1	AR-1016-2	4.760	4.690	4.683	5.272	4.891	4.859 E5	5.05
5) L1	AR-1016-3	6.728	6.650	6.629	7.258	7.472	6.947 E5	5.62

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ072718.M
 Title : GC EXTRACTABLES
 Last Update : Mon Jul 30 09:28:37 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ030901.D 1000 =PQ030899.D 750 =PQ030900.D
 250 =PQ030902.D 50 =PQ030903.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	4.261	4.346	4.271	4.892	4.515	4.457	E5	5.91
7)	L1 AR-1016-5	2.596	2.630	2.589	2.606	2.454	2.575	E5	2.70
8)	L2 AR-1221-1	2.286					2.286	E5	0.00
9)	L2 AR-1221-2	1.709					1.709	E5	0.00
10)	L2 AR-1221-3	5.209					5.209	E5	0.00
11)	L3 AR-1232-1	3.015					3.015	E5	0.00
12)	L3 AR-1232-2	3.423					3.423	E5	0.00
13)	L3 AR-1232-3	2.302					2.302	E5	0.00
14)	L3 AR-1232-4	1.349					1.349	E5	0.00
15)	L3 AR-1232-5	1.338					1.338	E5	0.00
16)	L4 AR-1242-1	3.082					3.082	E5	0.00
17)	L4 AR-1242-2	6.274					6.274	E5	0.00
18)	L4 AR-1242-3	3.309					3.309	E5	0.00
19)	L4 AR-1242-4	2.435					2.435	E5	0.00
20)	L4 AR-1242-5	3.519					3.519	E5	0.00
21)	L5 AR-1248-1	4.129					4.129	E5	0.00
22)	L5 AR-1248-2	4.375					4.375	E5	0.00
23)	L5 AR-1248-3	5.330					5.330	E5	0.00
24)	L5 AR-1248-4	7.860					7.860	E5	0.00
25)	L5 AR-1248-5	5.430					5.430	E5	0.00
26)	L6 AR-1254-1	8.006					8.006	E5	0.00
27)	L6 AR-1254-2	7.082					7.082	E5	0.00
28)	L6 AR-1254-3	1.173					1.173	E6	0.00
29)	L6 AR-1254-4	7.492					7.492	E5	0.00
30)	L6 AR-1254-5	1.045					1.045	E6	0.00
31)	L7 AR-1260-1	7.338	7.161	7.069	7.267	8.668	7.501	E5	8.81
32)	L7 AR-1260-2	0.928	0.886	0.872	0.923	1.201	0.962	E6	14.11
33)	L7 AR-1260-3	7.300	7.013	6.887	7.134	8.625	7.392	E5	9.55
34)	L7 AR-1260-4	1.732	1.692	1.651	1.677	1.821	1.714	E6	3.87
35)	L7 AR-1260-5	1.219	1.198	1.187	1.185	1.411	1.240	E6	7.79
36)	L8 AR-1262-1	6.429					6.429	E5	0.00
37)	L8 AR-1262-2	7.770					7.770	E5	0.00
38)	L8 AR-1262-3	1.077					1.077	E6	0.00
39)	L8 AR-1262-4	9.999					9.999	E5	0.00
40)	L8 AR-1262-5	2.102					2.102	E6	0.00
41)	L9 AR-1268-1	2.077					2.077	E6	0.00
42)	L9 AR-1268-2	1.968					1.968	E6	0.00
43)	L9 AR-1268-3	1.619					1.619	E6	0.00
44)	L9 AR-1268-4	6.971					6.971	E5	0.00
45)	L9 AR-1268-5	5.327					5.327	E6	0.00

(#) = Out of Range