

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072719\
 Data File : PQ041712.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Jul 2019 21:38
 Operator : SM\AJ
 Sample : AR1248ICC800
 Misc :
 ALS Vial : 76 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 29 00:51:05 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072719CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jul 29 00:49:22 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
1)	SA Tetrachlo...	5.701	4.759	99120534	81951984	38.160	39.162
2)	SA Decachlor...	12.098	10.513	705.4E6	402.3E6	75.659	75.154
Target Compounds							
21)	L5 AR-1248-1	7.144	6.210	52952475	48109627	755.676	748.516
22)	L5 AR-1248-2	7.459	6.500	65982505	63074342	732.855	746.111
23)	L5 AR-1248-3	7.687	6.550	82120467	64731566	745.527	745.481
24)	L5 AR-1248-4	8.134	6.753	111.9E6	80172563	737.119	754.589
25)	L5 AR-1248-5	8.175	7.208	107.8E6	95068408	730.558	748.078


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072719\
Data File : PQ041712.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Jul 2019 21:38
Operator : SM\AJ
Sample : AR1248ICC800
Misc :
ALS Vial : 76 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 29 00:51:05 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072719CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 29 00:49:22 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 99120534
Conc: 38.16 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 4.759 min
Delta R.T.: 0.000 min
Response: 81951984
Conc: 39.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 12.098 min
Delta R.T.: 0.000 min
Response: 705414973
Conc: 75.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.513 min
Delta R.T.: -0.002 min
Response: 402271563
Conc: 75.15 ng/ml




#21 AR-1248-1

R.T.: 7.144 min
Delta R.T.: 0.000 min
Response: 52952475
Conc: 755.68 ng/ml




#21 AR-1248-1

R.T.: 6.210 min
Delta R.T.: 0.000 min
Response: 48109627
Conc: 748.52 ng/ml




#22 AR-1248-2

R.T.: 7.459 min
Delta R.T.: 0.000 min
Response: 65982505
Conc: 732.86 ng/ml




#22 AR-1248-2

R.T.: 6.500 min
Delta R.T.: 0.000 min
Response: 63074342
Conc: 746.11 ng/ml




#23 AR-1248-3

R.T.: 7.687 min
Delta R.T.: 0.000 min
Response: 82120467
Conc: 745.53 ng/ml




#23 AR-1248-3

R.T.: 6.550 min
Delta R.T.: 0.000 min
Response: 64731566
Conc: 745.48 ng/ml




#24 AR-1248-4

R.T.: 8.134 min
Delta R.T.: 0.000 min
Response: 111861856
Conc: 737.12 ng/ml




#24 AR-1248-4

R.T.: 6.753 min
Delta R.T.: 0.000 min
Response: 80172563
Conc: 754.59 ng/ml



#25 AR-1248-5

R.T.: 8.175 min
Delta R.T.: 0.000 min
Response: 107784658
Conc: 730.56 ng/ml



#25 AR-1248-5

R.T.: 7.208 min
Delta R.T.: 0.000 min
Response: 95068408
Conc: 748.08 ng/ml