

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072921\
 Data File : PQ054142.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jul 2021 18:36
 Operator : AJ\MA
 Sample : M2940-07
 Misc : AR1660 LOD 25 PPB
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 04:14:32 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 30 02:37:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.204	4.208	408.4E6	237.7E6	17.490	16.813
2)	SA Decachlor...	11.379	9.545	319.9E6	259.9E6	19.011	20.603
Target Compounds							
3)	L1 AR-1016-1	6.531	5.472	18248894	11120608	24.897	24.566
4)	L1 AR-1016-2	6.555	5.492	27353096	18871836	24.352	25.460
5)	L1 AR-1016-3	6.622	5.685	18174616	9854593	26.348	27.335
6)	L1 AR-1016-4	6.730	5.735	14157577	7564915	24.886	26.505
7)	L1 AR-1016-5	7.041	5.966	13299659	9426752	24.570	25.873
8)	L2 AR-1221-1	5.427f	4.464	3233628	1061471	13.365	7.450 #
9)	L2 AR-1221-2	5.560	4.566	8605899	3797931	47.508	36.002
10)	L2 AR-1221-3	5.644	4.656	8900657	5558433	15.555	15.840
11)	L3 AR-1232-1	5.644	4.656	8900657	5558433	22.784	23.814
12)	L3 AR-1232-2	6.233	5.492	16041410	18871836	69.054	65.160
13)	L3 AR-1232-3	6.555	5.685	27353096	9854593	63.745	74.276
14)	L3 AR-1232-4	6.730	5.780	14157577	8889078	65.368	77.899
15)	L3 AR-1232-5	6.821	5.966	12334015	9426752	75.895	73.353
16)	L4 AR-1242-1	6.531	5.472	18248894	11120608	39.741	40.900
17)	L4 AR-1242-2	6.555	5.492	27353096	18871836	38.810	40.096
18)	L4 AR-1242-3	6.622	5.685	18174616	9854593	41.747	44.668
19)	L4 AR-1242-4	6.730	5.780	14157577	8889078	39.373	41.089
20)	L4 AR-1242-5	7.512	6.344	3929682	7925512	10.674	26.716 #
21)	L5 AR-1248-1	6.531	5.472	18248894	11120608	49.192	49.290
22)	L5 AR-1248-2	6.821	5.735	12334015	7564915	23.548	23.049
23)	L5 AR-1248-3	7.041	5.780	13299659	8889078	22.410	26.880
24)	L5 AR-1248-4	7.471	5.966	2930416	9426752	4.411	23.752 #
25)	L5 AR-1248-5	7.512	6.388	3929682	1274244	6.035	3.197 #
26)	L6 AR-1254-1	7.438	6.344	11473204	7925512	16.233	11.851 #
27)	L6 AR-1254-2	7.665	6.504	10497392	9496871	9.587	15.445 #
28)	L6 AR-1254-3	8.050	6.942f	5992060	15486384	5.095	16.039 #
29)	L6 AR-1254-4	8.345	7.166	5368991	3351541	6.092	5.029
30)	L6 AR-1254-5	8.772	7.591	32761468	27331063	33.573	28.022
31)	L7 AR-1260-1	8.215	7.060	26740246	23881138	27.235	31.860
32)	L7 AR-1260-2	8.478	7.257	29619875	30287024	26.088	28.995
33)	L7 AR-1260-3	8.844	7.412	16799123	26371577	23.749	30.302 #
34)	L7 AR-1260-4	9.086	7.895	22255745	15676133	25.158	25.177
35)	L7 AR-1260-5	9.431	8.142	39992869	41649568	23.755	24.309
36)	L8 AR-1262-1	8.844	7.591	16799123	27331063	16.984	60.476 #
37)	L8 AR-1262-2	9.431	8.142	39992869	41649568	22.156	22.117
38)	L8 AR-1262-3	9.791f	8.428	32988249	5402881	26.923	7.038 #
39)	L8 AR-1262-4	9.842f	8.491	16167493	27217851	17.859	21.771
40)	L8 AR-1262-5	10.573	8.993	11465971	10921205	17.299	18.923
41)	L9 AR-1268-1	0.000	8.428	0	5402881	N.D.	2.395 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072921\
 Data File : PQ054142.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jul 2021 18:36
 Operator : AJ\MA
 Sample : M2940-07
 Misc : AR1660 LOD 25 PPB
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOD-MDL-WATER-01-QT3-2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 04:14:32 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 30 02:37:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	9.842f	8.491	16167493	27217851	8.382	14.572 #
43) L9	AR-1268-3	0.000	8.700	0	2387724	N.D.	1.428 #
44) L9	AR-1268-4	10.573	8.993	11465971	10921205	15.676	16.747
45) L9	AR-1268-5	11.016	9.289	6430135	5385365	1.179	1.143



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072921\
Data File : PQ054142.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jul 2021 18:36
Operator : AJ\MA
Sample : M2940-07
Misc : AR1660 LOD 25 PPB
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 04:14:32 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072921.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 30 02:37:02 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.204 min
Delta R.T.: 0.000 min
Response: 408436848
Conc: 17.49 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#1 Tetrachloro-m-xylene

R.T.: 4.208 min
Delta R.T.: 0.002 min
Response: 237695167
Conc: 16.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.379 min
Delta R.T.: 0.000 min
Response: 319878359
Conc: 19.01 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.545 min
Delta R.T.: -0.003 min
Response: 259887316
Conc: 20.60 ng/ml



#3 AR-1016-1


R.T.: 6.531 min
Delta R.T.: 0.003 min
Response: 18248894
Conc: 24.90 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#3 AR-1016-1

R.T.: 5.472 min
Delta R.T.: 0.001 min
Response: 11120608
Conc: 24.57 ng/ml




#4 AR-1016-2

R.T.: 6.555 min
Delta R.T.: 0.002 min
Response: 27353096
Conc: 24.35 ng/ml



#4 AR-1016-2


R.T.: 5.492 min
Delta R.T.: 0.002 min
Response: 18871836
Conc: 25.46 ng/ml



#5 AR-1016-3


R.T.: 6.622 min
Delta R.T.: 0.003 min
Response: 18174616
Conc: 26.35 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#5 AR-1016-3

R.T.: 5.685 min
Delta R.T.: 0.001 min
Response: 9854593
Conc: 27.34 ng/ml




#6 AR-1016-4

R.T.: 6.730 min
Delta R.T.: 0.002 min
Response: 14157577
Conc: 24.89 ng/ml



#6 AR-1016-4


R.T.: 5.735 min
Delta R.T.: 0.000 min
Response: 7564915
Conc: 26.50 ng/ml



#7 AR-1016-5


R.T.: 7.041 min
Delta R.T.: 0.000 min
Response: 13299659
Conc: 24.57 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#7 AR-1016-5

R.T.: 5.966 min
Delta R.T.: 0.000 min
Response: 9426752
Conc: 25.87 ng/ml




#8 AR-1221-1

R.T.: 5.427 min
Delta R.T.: -0.026 min
Response: 3233628
Conc: 13.37 ng/ml



#8 AR-1221-1


R.T.: 4.464 min
Delta R.T.: -0.001 min
Response: 1061471
Conc: 7.45 ng/ml



#9 AR-1221-2


R.T.: 5.560 min
Delta R.T.: 0.006 min
Response: 8605899
Conc: 47.51 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#9 AR-1221-2

R.T.: 4.566 min
Delta R.T.: 0.000 min
Response: 3797931
Conc: 36.00 ng/ml




#10 AR-1221-3

R.T.: 5.644 min
Delta R.T.: 0.003 min
Response: 8900657
Conc: 15.56 ng/ml



#10 AR-1221-3


R.T.: 4.656 min
Delta R.T.: 0.001 min
Response: 5558433
Conc: 15.84 ng/ml



#11 AR-1232-1


R.T.: 5.644 min
Delta R.T.: 0.002 min
Response: 8900657
Conc: 22.78 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#11 AR-1232-1

R.T.: 4.656 min
Delta R.T.: 0.001 min
Response: 5558433
Conc: 23.81 ng/ml




#12 AR-1232-2

R.T.: 6.233 min
Delta R.T.: 0.000 min
Response: 16041410
Conc: 69.05 ng/ml



#12 AR-1232-2


R.T.: 5.492 min
Delta R.T.: 0.000 min
Response: 18871836
Conc: 65.16 ng/ml



#13 AR-1232-3


R.T.: 6.555 min
Delta R.T.: 0.001 min
Response: 27353096
Conc: 63.74 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#13 AR-1232-3

R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 9854593
Conc: 74.28 ng/ml




#14 AR-1232-4

R.T.: 6.730 min
Delta R.T.: 0.000 min
Response: 14157577
Conc: 65.37 ng/ml



#14 AR-1232-4


R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 8889078
Conc: 77.90 ng/ml



#15 AR-1232-5


R.T.: 6.821 min
Delta R.T.: 0.000 min
Response: 12334015
Conc: 75.90 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#15 AR-1232-5

R.T.: 5.966 min
Delta R.T.: 0.000 min
Response: 9426752
Conc: 73.35 ng/ml




#16 AR-1242-1

R.T.: 6.531 min
Delta R.T.: 0.002 min
Response: 18248894
Conc: 39.74 ng/ml



#16 AR-1242-1


R.T.: 5.472 min
Delta R.T.: 0.000 min
Response: 11120608
Conc: 40.90 ng/ml



#17 AR-1242-2


R.T.: 6.555 min
Delta R.T.: 0.000 min
Response: 27353096
Conc: 38.81 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#17 AR-1242-2

R.T.: 5.492 min
Delta R.T.: 0.000 min
Response: 18871836
Conc: 40.10 ng/ml




#18 AR-1242-3

R.T.: 6.622 min
Delta R.T.: 0.001 min
Response: 18174616
Conc: 41.75 ng/ml



#18 AR-1242-3


R.T.: 5.685 min
Delta R.T.: 0.000 min
Response: 9854593
Conc: 44.67 ng/ml



#19 AR-1242-4


R.T.: 6.730 min
Delta R.T.: 0.000 min
Response: 14157577
Conc: 39.37 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#19 AR-1242-4

R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 8889078
Conc: 41.09 ng/ml




#20 AR-1242-5

R.T.: 7.512 min
Delta R.T.: 0.001 min
Response: 3929682
Conc: 10.67 ng/ml



#20 AR-1242-5


R.T.: 6.344 min
Delta R.T.: -0.001 min
Response: 7925512
Conc: 26.72 ng/ml



#21 AR-1248-1


R.T.: 6.531 min
Delta R.T.: 0.002 min
Response: 18248894
Conc: 49.19 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#21 AR-1248-1

R.T.: 5.472 min
Delta R.T.: 0.000 min
Response: 11120608
Conc: 49.29 ng/ml




#22 AR-1248-2

R.T.: 6.821 min
Delta R.T.: 0.000 min
Response: 12334015
Conc: 23.55 ng/ml



#22 AR-1248-2


R.T.: 5.735 min
Delta R.T.: 0.000 min
Response: 7564915
Conc: 23.05 ng/ml



#23 AR-1248-3


R.T.: 7.041 min
Delta R.T.: 0.000 min
Response: 13299659
Conc: 22.41 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#23 AR-1248-3

R.T.: 5.780 min
Delta R.T.: 0.000 min
Response: 8889078
Conc: 26.88 ng/ml




#24 AR-1248-4

R.T.: 7.471 min
Delta R.T.: 0.002 min
Response: 2930416
Conc: 4.41 ng/ml



#24 AR-1248-4


R.T.: 5.966 min
Delta R.T.: 0.000 min
Response: 9426752
Conc: 23.75 ng/ml



#25 AR-1248-5


R.T.: 7.512 min
Delta R.T.: 0.002 min
Response: 3929682
Conc: 6.04 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#25 AR-1248-5

R.T.: 6.388 min
Delta R.T.: 0.000 min
Response: 1274244
Conc: 3.20 ng/ml




#26 AR-1254-1

R.T.: 7.438 min
Delta R.T.: 0.000 min
Response: 11473204
Conc: 16.23 ng/ml



#26 AR-1254-1


R.T.: 6.344 min
Delta R.T.: 0.000 min
Response: 7925512
Conc: 11.85 ng/ml



#27 AR-1254-2


R.T.: 7.665 min
Delta R.T.: -0.001 min
Response: 10497392
Conc: 9.59 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#27 AR-1254-2

R.T.: 6.504 min
Delta R.T.: 0.000 min
Response: 9496871
Conc: 15.45 ng/ml




#28 AR-1254-3

R.T.: 8.050 min
Delta R.T.: 0.000 min
Response: 5992060
Conc: 5.09 ng/ml



#28 AR-1254-3


R.T.: 6.942 min
Delta R.T.: 0.017 min
Response: 15486384
Conc: 16.04 ng/ml



#29 AR-1254-4


R.T.: 8.345 min
Delta R.T.: -0.002 min
Response: 5368991
Conc: 6.09 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#29 AR-1254-4

R.T.: 7.166 min
Delta R.T.: 0.002 min
Response: 3351541
Conc: 5.03 ng/ml




#30 AR-1254-5

R.T.: 8.772 min
Delta R.T.: -0.002 min
Response: 32761468
Conc: 33.57 ng/ml



#30 AR-1254-5


R.T.: 7.591 min
Delta R.T.: -0.001 min
Response: 27331063
Conc: 28.02 ng/ml



#31 AR-1260-1


R.T.: 8.215 min
Delta R.T.: 0.000 min
Response: 26740246
Conc: 27.23 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#31 AR-1260-1

R.T.: 7.060 min
Delta R.T.: 0.000 min
Response: 23881138
Conc: 31.86 ng/ml




#32 AR-1260-2

R.T.: 8.478 min
Delta R.T.: 0.000 min
Response: 29619875
Conc: 26.09 ng/ml



#32 AR-1260-2


R.T.: 7.257 min
Delta R.T.: 0.000 min
Response: 30287024
Conc: 28.99 ng/ml



#33 AR-1260-3


R.T.: 8.844 min
Delta R.T.: -0.001 min
Response: 16799123
Conc: 23.75 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#33 AR-1260-3

R.T.: 7.412 min
Delta R.T.: 0.000 min
Response: 26371577
Conc: 30.30 ng/ml




#34 AR-1260-4

R.T.: 9.086 min
Delta R.T.: -0.001 min
Response: 22255745
Conc: 25.16 ng/ml



#34 AR-1260-4


R.T.: 7.895 min
Delta R.T.: 0.000 min
Response: 15676133
Conc: 25.18 ng/ml



#35 AR-1260-5


R.T.: 9.431 min
Delta R.T.: 0.000 min
Response: 39992869
Conc: 23.76 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#35 AR-1260-5

R.T.: 8.142 min
Delta R.T.: 0.000 min
Response: 41649568
Conc: 24.31 ng/ml




#36 AR-1262-1

R.T.: 8.844 min
Delta R.T.: -0.002 min
Response: 16799123
Conc: 16.98 ng/ml



#36 AR-1262-1


R.T.: 7.591 min
Delta R.T.: 0.000 min
Response: 27331063
Conc: 60.48 ng/ml



#37 AR-1262-2


R.T.: 9.431 min
Delta R.T.: 0.000 min
Response: 39992869
Conc: 22.16 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#37 AR-1262-2

R.T.: 8.142 min
Delta R.T.: 0.000 min
Response: 41649568
Conc: 22.12 ng/ml




#38 AR-1262-3

R.T.: 9.791 min
Delta R.T.: 0.028 min
Response: 32988249
Conc: 26.92 ng/ml



#38 AR-1262-3


R.T.: 8.428 min
Delta R.T.: -0.001 min
Response: 5402881
Conc: 7.04 ng/ml



#39 AR-1262-4


R.T.: 9.842 min
Delta R.T.: -0.018 min
Response: 16167493
Conc: 17.86 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#39 AR-1262-4

R.T.: 8.491 min
Delta R.T.: -0.003 min
Response: 27217851
Conc: 21.77 ng/ml




#40 AR-1262-5

R.T.: 10.573 min
Delta R.T.: -0.001 min
Response: 11465971
Conc: 17.30 ng/ml



#40 AR-1262-5


R.T.: 8.993 min
Delta R.T.: -0.002 min
Response: 10921205
Conc: 18.92 ng/ml



#41 AR-1268-1


R.T.: 0.000 min
Exp R.T.: 9.760 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#41 AR-1268-1

R.T.: 8.428 min
Delta R.T.: 0.000 min
Response: 5402881
Conc: 2.40 ng/ml




#42 AR-1268-2

R.T.: 9.842 min
Delta R.T.: -0.022 min
Response: 16167493
Conc: 8.38 ng/ml



#42 AR-1268-2


R.T.: 8.491 min
Delta R.T.: -0.004 min
Response: 27217851
Conc: 14.57 ng/ml



#43 AR-1268-3


R.T.: 0.000 min
Exp R.T.: 10.105 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021




#43 AR-1268-3

R.T.: 8.700 min
Delta R.T.: -0.002 min
Response: 2387724
Conc: 1.43 ng/ml




#44 AR-1268-4

R.T.: 10.573 min
Delta R.T.: 0.000 min
Response: 11465971
Conc: 15.68 ng/ml



#44 AR-1268-4


R.T.: 8.993 min
Delta R.T.: -0.001 min
Response: 10921205
Conc: 16.75 ng/ml



#45 AR-1268-5

R.T.: 11.016 min
Delta R.T.: 0.000 min
Response: 6430135
Conc: 1.18 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOD-MDL-WATER-01-QT3-2021



#45 AR-1268-5

R.T.: 9.289 min
Delta R.T.: -0.002 min
Response: 5385365
Conc: 1.14 ng/ml