

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072921\
 Data File : PQ054169.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jul 2021 02:33
 Operator : AJ\MA
 Sample : M2940-08
 Misc : AR1254 LOQ 50 PPB
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOQ-WATER-02-QT3-2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 06:41:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 30 02:37:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.203	4.207	432.6E6	243.3E6	18.523	17.210
2)	SA Decachlor...	11.377	9.544	308.3E6	243.6E6	18.324	19.308
Target Compounds							
3)	L1 AR-1016-1	6.499f	5.494f	5298817	2602660	7.229	5.749
4)	L1 AR-1016-2	0.000	5.494	0	2602660	N.D.	3.511 #
5)	L1 AR-1016-3	0.000	5.689	0	522418	N.D.	1.449 #
6)	L1 AR-1016-4	0.000	5.733	0	15801121	N.D.	55.362 #
7)	L1 AR-1016-5	7.041	5.964	13165587	8054095	24.323	22.105
8)	L2 AR-1221-1	5.426f	0.000	1434872	0	5.931	N.D. #
9)	L2 AR-1221-2	5.562	4.564	7845425	2699623	43.310	25.591 #
10)	L2 AR-1221-3	0.000	4.641	0	1345430	N.D.	3.834 #
11)	L3 AR-1232-1	0.000	4.641	0	1345430	N.D.	5.764 #
12)	L3 AR-1232-2	0.000	5.494	0	2602660	N.D.	8.986 #
13)	L3 AR-1232-3	0.000	5.689	0	522418	N.D.	3.938 #
14)	L3 AR-1232-4	0.000	5.778	0	3686562	N.D.	32.307 #
15)	L3 AR-1232-5	6.821	5.964	21517995	8054095	132.407	62.672 #
16)	L4 AR-1242-1	0.000	5.494f	0	2602660	N.D.	9.572 #
17)	L4 AR-1242-2	0.000	5.494	0	2602660	N.D.	5.530 #
18)	L4 AR-1242-3	0.000	5.689	0	522418	N.D.	2.368 #
19)	L4 AR-1242-4	0.000	5.778	0	3686562	N.D.	17.041 #
20)	L4 AR-1242-5	7.512	6.344	11775064	38093369	31.985	128.411 #
21)	L5 AR-1248-1	0.000	5.494f	0	2602660	N.D.	11.536 #
22)	L5 AR-1248-2	6.821	5.733	21517995	15801121	41.081	48.144
23)	L5 AR-1248-3	7.041	5.778	13165587	3686562	22.184	11.148 #
24)	L5 AR-1248-4	7.471	5.964	17330266	8054095	26.089	20.294
25)	L5 AR-1248-5	7.512	6.388	11775064	4719148	18.085	11.841 #
26)	L6 AR-1254-1	7.438	6.344	39761584	38093369	56.258	56.963
27)	L6 AR-1254-2	7.666	6.503	60471178	35193153	55.226	57.237
28)	L6 AR-1254-3	8.050	6.924	63348125	51883134	53.864	53.733
29)	L6 AR-1254-4	8.346	7.164	46769465	36764121	53.067	55.164
30)	L6 AR-1254-5	8.773	7.592	57853995	52179492	59.288	53.499
31)	L7 AR-1260-1	8.214	7.060	26771455	28341372	27.267	37.810 #
32)	L7 AR-1260-2	8.478	7.256	30609511	28015697	26.960	26.821
33)	L7 AR-1260-3	8.835	7.410	5843226	27658285	8.260	31.781 #
34)	L7 AR-1260-4	9.081	7.894	18957378	2739872	21.430	4.400 #
35)	L7 AR-1260-5	9.429	8.140	7730557	6715696	4.592	3.920
36)	L8 AR-1262-1	8.835	7.592	5843226	52179492	5.907	115.459 #
37)	L8 AR-1262-2	9.429	8.140	7730557	6715696	4.283	3.566
38)	L8 AR-1262-3	9.792f	0.000	6393711	0	5.218	N.D. #
39)	L8 AR-1262-4	0.000	8.489	0	3992428	N.D.	3.194 #
40)	L8 AR-1262-5	0.000	8.995	0	655553	N.D.	1.136 #
42)	L9 AR-1268-2	0.000	8.489	0	3992428	N.D.	2.137 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072921\
 Data File : PQ054169.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jul 2021 02:33
 Operator : AJ\MA
 Sample : M2940-08
 Misc : AR1254 LOQ 50 PPB
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 LOQ-WATER-02-QT3-2021

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 06:41:11 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072921.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jul 30 02:37:02 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
43) L9	AR-1268-3	0.000	8.699	0	1317161	N.D.	0.788 #
44) L9	AR-1268-4	0.000	8.995	0	655553	N.D.	1.005 #
45) L9	AR-1268-5	0.000	9.286	0	2189274	N.D.	0.465 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ072921\
Data File : PQ054169.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jul 2021 02:33
Operator : AJ\MA
Sample : M2940-08
Misc : AR1254 LOQ 50 PPB
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 06:41:11 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072921.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jul 30 02:37:02 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 5.203 min
Delta R.T.: 0.000 min
Response: 432562914
Conc: 18.52 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#1 Tetrachloro-m-xylene

R.T.: 4.207 min
Delta R.T.: 0.000 min
Response: 243304793
Conc: 17.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 11.377 min
Delta R.T.: -0.002 min
Response: 308314273
Conc: 18.32 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.544 min
Delta R.T.: -0.004 min
Response: 243553808
Conc: 19.31 ng/ml



#3 AR-1016-1


R.T.: 6.499 min
Delta R.T.: -0.029 min
Response: 5298817
Conc: 7.23 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#3 AR-1016-1

R.T.: 5.494 min
Delta R.T.: 0.023 min
Response: 2602660
Conc: 5.75 ng/ml




#4 AR-1016-2

R.T.: 0.000 min
Exp R.T. : 6.553 min
Response: 0
Conc: N.D.



#4 AR-1016-2


R.T.: 5.494 min
Delta R.T.: 0.003 min
Response: 2602660
Conc: 3.51 ng/ml



#5 AR-1016-3


R.T.: 0.000 min
Exp R.T.: 6.619 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#5 AR-1016-3

R.T.: 5.689 min
Delta R.T.: 0.005 min
Response: 522418
Conc: 1.45 ng/ml




#6 AR-1016-4

R.T.: 0.000 min
Exp R.T.: 6.728 min
Response: 0
Conc: N.D.



#6 AR-1016-4


R.T.: 5.733 min
Delta R.T.: -0.001 min
Response: 15801121
Conc: 55.36 ng/ml



#7 AR-1016-5


R.T.: 7.041 min
Delta R.T.: 0.000 min
Response: 13165587
Conc: 24.32 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#7 AR-1016-5

R.T.: 5.964 min
Delta R.T.: 0.000 min
Response: 8054095
Conc: 22.11 ng/ml




#8 AR-1221-1

R.T.: 5.426 min
Delta R.T.: -0.028 min
Response: 1434872
Conc: 5.93 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 4.465 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 5.562 min
Delta R.T.: 0.008 min
Response: 7845425
Conc: 43.31 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#9 AR-1221-2

R.T.: 4.564 min
Delta R.T.: -0.002 min
Response: 2699623
Conc: 25.59 ng/ml




#10 AR-1221-3

R.T.: 0.000 min
Exp R.T. : 5.641 min
Response: 0
Conc: N.D.



#10 AR-1221-3


R.T.: 4.641 min
Delta R.T.: -0.014 min
Response: 1345430
Conc: 3.83 ng/ml



#11 AR-1232-1


R.T.: 0.000 min
Exp R.T.: 5.642 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#11 AR-1232-1

R.T.: 4.641 min
Delta R.T.: -0.014 min
Response: 1345430
Conc: 5.76 ng/ml




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T.: 6.233 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 5.494 min
Delta R.T.: 0.003 min
Response: 2602660
Conc: 8.99 ng/ml



#13 AR-1232-3


R.T.: 0.000 min
Exp R.T.: 6.554 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#13 AR-1232-3

R.T.: 5.689 min
Delta R.T.: 0.004 min
Response: 522418
Conc: 3.94 ng/ml




#14 AR-1232-4

R.T.: 0.000 min
Exp R.T.: 6.729 min
Response: 0
Conc: N.D.



#14 AR-1232-4


R.T.: 5.778 min
Delta R.T.: -0.002 min
Response: 3686562
Conc: 32.31 ng/ml



#15 AR-1232-5


R.T.: 6.821 min
Delta R.T.: 0.000 min
Response: 21517995
Conc: 132.41 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#15 AR-1232-5

R.T.: 5.964 min
Delta R.T.: -0.002 min
Response: 8054095
Conc: 62.67 ng/ml




#16 AR-1242-1

R.T.: 0.000 min
Exp R.T. : 6.530 min
Response: 0
Conc: N.D.



#16 AR-1242-1


R.T.: 5.494 min
Delta R.T.: 0.022 min
Response: 2602660
Conc: 9.57 ng/ml



#17 AR-1242-2


R.T.: 0.000 min
Exp R.T.: 6.554 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#17 AR-1242-2

R.T.: 5.494 min
Delta R.T.: 0.002 min
Response: 2602660
Conc: 5.53 ng/ml




#18 AR-1242-3

R.T.: 0.000 min
Exp R.T.: 6.620 min
Response: 0
Conc: N.D.



#18 AR-1242-3


R.T.: 5.689 min
Delta R.T.: 0.004 min
Response: 522418
Conc: 2.37 ng/ml



#19 AR-1242-4


R.T.: 0.000 min
Exp R.T.: 6.729 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#19 AR-1242-4

R.T.: 5.778 min
Delta R.T.: -0.003 min
Response: 3686562
Conc: 17.04 ng/ml




#20 AR-1242-5

R.T.: 7.512 min
Delta R.T.: 0.002 min
Response: 11775064
Conc: 31.98 ng/ml



#20 AR-1242-5


R.T.: 6.344 min
Delta R.T.: -0.002 min
Response: 38093369
Conc: 128.41 ng/ml



#21 AR-1248-1


R.T.: 0.000 min
Exp R.T.: 6.529 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#21 AR-1248-1

R.T.: 5.494 min
Delta R.T.: 0.022 min
Response: 2602660
Conc: 11.54 ng/ml




#22 AR-1248-2

R.T.: 6.821 min
Delta R.T.: 0.000 min
Response: 21517995
Conc: 41.08 ng/ml



#22 AR-1248-2


R.T.: 5.733 min
Delta R.T.: -0.002 min
Response: 15801121
Conc: 48.14 ng/ml



#23 AR-1248-3


R.T.: 7.041 min
Delta R.T.: 0.000 min
Response: 13165587
Conc: 22.18 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#23 AR-1248-3

R.T.: 5.778 min
Delta R.T.: -0.002 min
Response: 3686562
Conc: 11.15 ng/ml




#24 AR-1248-4

R.T.: 7.471 min
Delta R.T.: 0.000 min
Response: 17330266
Conc: 26.09 ng/ml



#24 AR-1248-4


R.T.: 5.964 min
Delta R.T.: -0.001 min
Response: 8054095
Conc: 20.29 ng/ml



#25 AR-1248-5


R.T.: 7.512 min
Delta R.T.: 0.002 min
Response: 11775064
Conc: 18.08 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#25 AR-1248-5

R.T.: 6.388 min
Delta R.T.: 0.000 min
Response: 4719148
Conc: 11.84 ng/ml




#26 AR-1254-1

R.T.: 7.438 min
Delta R.T.: 0.000 min
Response: 39761584
Conc: 56.26 ng/ml



#26 AR-1254-1


R.T.: 6.344 min
Delta R.T.: 0.000 min
Response: 38093369
Conc: 56.96 ng/ml



#27 AR-1254-2


R.T.: 7.666 min
Delta R.T.: 0.000 min
Response: 60471178
Conc: 55.23 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#27 AR-1254-2

R.T.: 6.503 min
Delta R.T.: 0.000 min
Response: 35193153
Conc: 57.24 ng/ml




#28 AR-1254-3

R.T.: 8.050 min
Delta R.T.: 0.000 min
Response: 63348125
Conc: 53.86 ng/ml



#28 AR-1254-3


R.T.: 6.924 min
Delta R.T.: 0.000 min
Response: 51883134
Conc: 53.73 ng/ml



#29 AR-1254-4


R.T.: 8.346 min
Delta R.T.: 0.000 min
Response: 46769465
Conc: 53.07 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#29 AR-1254-4

R.T.: 7.164 min
Delta R.T.: 0.000 min
Response: 36764121
Conc: 55.16 ng/ml




#30 AR-1254-5

R.T.: 8.773 min
Delta R.T.: -0.001 min
Response: 57853995
Conc: 59.29 ng/ml



#30 AR-1254-5


R.T.: 7.592 min
Delta R.T.: 0.000 min
Response: 52179492
Conc: 53.50 ng/ml



#31 AR-1260-1


R.T.: 8.214 min
Delta R.T.: 0.000 min
Response: 26771455
Conc: 27.27 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#31 AR-1260-1

R.T.: 7.060 min
Delta R.T.: 0.000 min
Response: 28341372
Conc: 37.81 ng/ml




#32 AR-1260-2

R.T.: 8.478 min
Delta R.T.: 0.000 min
Response: 30609511
Conc: 26.96 ng/ml



#32 AR-1260-2


R.T.: 7.256 min
Delta R.T.: -0.001 min
Response: 28015697
Conc: 26.82 ng/ml



#33 AR-1260-3


R.T.: 8.835 min
Delta R.T.: -0.010 min
Response: 5843226
Conc: 8.26 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#33 AR-1260-3

R.T.: 7.410 min
Delta R.T.: -0.003 min
Response: 27658285
Conc: 31.78 ng/ml




#34 AR-1260-4

R.T.: 9.081 min
Delta R.T.: -0.007 min
Response: 18957378
Conc: 21.43 ng/ml



#34 AR-1260-4


R.T.: 7.894 min
Delta R.T.: -0.002 min
Response: 2739872
Conc: 4.40 ng/ml



#35 AR-1260-5


R.T.: 9.429 min
Delta R.T.: -0.003 min
Response: 7730557
Conc: 4.59 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#35 AR-1260-5

R.T.: 8.140 min
Delta R.T.: -0.003 min
Response: 6715696
Conc: 3.92 ng/ml




#36 AR-1262-1

R.T.: 8.835 min
Delta R.T.: -0.010 min
Response: 5843226
Conc: 5.91 ng/ml



#36 AR-1262-1


R.T.: 7.592 min
Delta R.T.: 0.000 min
Response: 52179492
Conc: 115.46 ng/ml



#37 AR-1262-2


R.T.: 9.429 min
Delta R.T.: -0.003 min
Response: 7730557
Conc: 4.28 ng/ml

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#37 AR-1262-2

R.T.: 8.140 min
Delta R.T.: -0.002 min
Response: 6715696
Conc: 3.57 ng/ml




#38 AR-1262-3

R.T.: 9.792 min
Delta R.T.: 0.029 min
Response: 6393711
Conc: 5.22 ng/ml



#38 AR-1262-3


R.T.: 0.000 min
Exp R.T.: 8.429 min
Response: 0
Conc: N.D.



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T.: 9.861 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#39 AR-1262-4

R.T.: 8.489 min
Delta R.T.: -0.005 min
Response: 3992428
Conc: 3.19 ng/ml




#40 AR-1262-5

R.T.: 0.000 min
Exp R.T.: 10.574 min
Response: 0
Conc: N.D.



#40 AR-1262-5


R.T.: 8.995 min
Delta R.T.: 0.000 min
Response: 655553
Conc: 1.14 ng/ml



#42 AR-1268-2


R.T.: 0.000 min
Exp R.T.: 9.864 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#42 AR-1268-2

R.T.: 8.489 min
Delta R.T.: -0.006 min
Response: 3992428
Conc: 2.14 ng/ml




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T.: 10.105 min
Response: 0
Conc: N.D.



#43 AR-1268-3


R.T.: 8.699 min
Delta R.T.: -0.003 min
Response: 1317161
Conc: 0.79 ng/ml



#44 AR-1268-4


R.T.: 0.000 min
Exp R.T.: 10.573 min
Response: 0
Conc: N.D.

Instrument :
ECD_Q
ClientSampleId :
LOQ-WATER-02-QT3-2021




#44 AR-1268-4

R.T.: 8.995 min
Delta R.T.: 0.001 min
Response: 655553
Conc: 1.01 ng/ml



#45 AR-1268-5

R.T.: 0.000 min
Exp R.T.: 11.016 min
Response: 0
Conc: N.D.



#45 AR-1268-5

R.T.: 9.286 min
Delta R.T.: -0.004 min
Response: 2189274
Conc: 0.46 ng/ml