Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080219\
Data File : PQ041975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2019 11:39
Operator : SM\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254331

Manual Integrations APPROVED
Ankita
8/5/2019 2:42:04 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 00:09:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072719CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 29 04:03:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
	System Monitoring Compounds						
1)	SA Tetrachlo...	5.700	4.752	48951127	43153848	21.344	22.392
2)	SA Decachlor...	12.095	10.508	262.6E6	177.7E6	31.141	36.744
	Target Compounds						
26)	L6 AR-1254-1	8.104	7.156	62394031	84037985	448.508	459.514m
27)	L6 AR-1254-2	8.341	7.324	96824788	73487012	423.032	453.984m
28)	L6 AR-1254-3	8.735	7.771	119.2E6	128.6E6	435.954	445.905m
29)	L6 AR-1254-4	9.037	8.021	94810339	93954417	385.134	432.825
30)	L6 AR-1254-5	9.475	8.467	101.8E6	125.5E6	375.807	417.227

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Quantitation Report (QT Reviewed)

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080219\
Data File : PQ041975.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2019 11:39
Operator : SM\AJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254331

Manual Integrations APPROVED
Ankita
8/5/2019 2:42:04 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 00:09:27 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072719CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 29 04:03:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm