

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080224\
 Data File : PQ067604.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Aug 2024 11:04
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 16:08:07 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 24 19:21:12 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.460	2.828	315.0E6	333.7E6	55.296	55.117
2)	SA Decachlor...	8.739	7.612	208.8E6	290.7E6	54.536	50.564
Target Compounds							
26)	L6 AR-1254-1	5.408	4.596	114.0E6	167.0E6	518.637	493.992
27)	L6 AR-1254-2	5.629	4.743	170.9E6	151.7E6	513.203	496.150
28)	L6 AR-1254-3	5.988	5.126	172.7E6	229.1E6	506.347	484.706
29)	L6 AR-1254-4	6.280	5.354	112.8E6	119.2E6	497.463	449.751
30)	L6 AR-1254-5	6.699	5.760	142.7E6	211.1E6	507.501m	490.433

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080224\
Data File : PQ067604.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2024 11:04
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 16:08:07 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 24 19:21:12 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

