

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080224\
 Data File : PQ067619.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Aug 2024 20:03
 Operator : YP\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 02 23:49:26 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072424.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jul 24 19:21:12 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.459	2.828	320.0E6	337.9E6	56.171	55.796
2)	SA Decachlor...	8.730	7.610	217.0E6	301.0E6	56.678	52.363
Target Compounds							
26)	L6 AR-1254-1	5.405	4.596	115.6E6	174.4E6	526.256	515.847
27)	L6 AR-1254-2	5.626	4.743	174.4E6	157.2E6	523.718	514.111
28)	L6 AR-1254-3	5.985	5.126	180.9E6	241.1E6	530.252	510.114
29)	L6 AR-1254-4	6.276	5.353	122.4E6	138.0E6	539.942	520.654
30)	L6 AR-1254-5	6.695	5.759	143.7E6	224.4E6	510.935m	521.316

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080224\
Data File : PQ067619.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Aug 2024 20:03
Operator : YP\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 02 23:49:26 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072424.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jul 24 19:21:12 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

