

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080422\
 Data File : PQ058747.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Aug 2022 03:04
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 ICVPQ080422AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 05 03:40:05 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080422.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 05 03:39:19 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.687	4.049	346.4E6	283.9E6	50.662	51.862
2) SA Decachlor...	10.592	9.166	389.4E6	220.9E6	52.145	51.852
Target Compounds						
21) L5 AR-1248-1	5.867	5.147	51088751	52705287	501.934	521.124
22) L5 AR-1248-2	6.139	5.384	71384384	74449586	498.097	516.297
23) L5 AR-1248-3	6.346	5.427	89014750	77153490	518.347	513.194
24) L5 AR-1248-4	6.752	5.602	94003912	92216133	512.206	528.032
25) L5 AR-1248-5	6.791	5.996	99711386	86197572	512.188	520.834

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080422\
Data File : PQ058747.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Aug 2022 03:04
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
ICVPQ080422AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 05 03:40:05 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080422.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 05 03:39:19 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

