

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ080519.M
 Title : GC EXTRACTABLES
 Last Update : Tue Aug 06 01:47:44 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ041998.D 1000 =PQ041996.D 750 =PQ041997.D
 250 =PQ041999.D 50 =PQ042000.D

Compound			500	1000	750	250	50	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	2.865	3.031	2.967	3.322	3.167	3.070	E6	5.80
2)	SA	Decachlorobiphenyl	0.966	0.925	0.945	1.011	1.072	0.984	E7	5.98
3)	L1	AR-1016-1	1.380	1.343	1.338	1.581	1.665	1.462	E5	10.36
4)	L1	AR-1016-2	2.046	2.012	2.010	2.321	2.404	2.159	E5	8.74
5)	L1	AR-1016-3	1.273	1.229	1.236	1.442	1.417	1.319	E5	7.75
6)	L1	AR-1016-4	1.053	1.037	1.036	1.199	1.234	1.112	E5	8.68
7)	L1	AR-1016-5	1.049	1.017	1.007	1.198	1.235	1.101	E5	9.75
8)	L2	AR-1221-1	3.203					3.203	E4	0.00
9)	L2	AR-1221-2	2.424					2.424	E4	0.00
10)	L2	AR-1221-3	8.080					8.080	E4	0.00
11)	L3	AR-1232-1	6.751					6.751	E4	0.00
12)	L3	AR-1232-2	3.785					3.785	E4	0.00
13)	L3	AR-1232-3	8.381					8.381	E4	0.00
14)	L3	AR-1232-4	4.337					4.337	E4	0.00
15)	L3	AR-1232-5	2.722					2.722	E4	0.00
16)	L4	AR-1242-1	1.125					1.125	E5	0.00
17)	L4	AR-1242-2	1.637					1.637	E5	0.00
18)	L4	AR-1242-3	1.020					1.020	E5	0.00
19)	L4	AR-1242-4	8.549					8.549	E4	0.00
20)	L4	AR-1242-5	1.153					1.153	E5	0.00
21)	L5	AR-1248-1	8.046					8.046	E4	0.00
22)	L5	AR-1248-2	1.035					1.035	E5	0.00
23)	L5	AR-1248-3	1.322					1.322	E5	0.00
24)	L5	AR-1248-4	1.820					1.820	E5	0.00
25)	L5	AR-1248-5	1.811					1.811	E5	0.00
26)	L6	AR-1254-1	1.886					1.886	E5	0.00
27)	L6	AR-1254-2	3.195					3.195	E5	0.00
28)	L6	AR-1254-3	3.913					3.913	E5	0.00
29)	L6	AR-1254-4	3.263					3.263	E5	0.00
30)	L6	AR-1254-5	3.653					3.653	E5	0.00
31)	L7	AR-1260-1	2.419	2.308	2.311	2.652	2.799	2.497	E5	8.76
32)	L7	AR-1260-2	3.330	3.186	3.187	3.628	3.952	3.457	E5	9.56
33)	L7	AR-1260-3	2.900	2.789	2.809	3.128	3.252	2.975	E5	6.89
34)	L7	AR-1260-4	3.365	3.209	3.232	3.557	3.813	3.435	E5	7.35
35)	L7	AR-1260-5	8.161	8.006	8.029	8.554	8.815	8.313	E5	4.29
36)	L8	AR-1262-1	3.876					3.876	E5	0.00
37)	L8	AR-1262-2	8.167					8.167	E5	0.00
38)	L8	AR-1262-3	5.874					5.874	E5	0.00
39)	L8	AR-1262-4	4.705					4.705	E5	0.00
40)	L8	AR-1262-5	3.507					3.507	E5	0.00
41)	L9	AR-1268-1	1.447					1.447	E6	0.00
42)	L9	AR-1268-2	1.359					1.359	E6	0.00
43)	L9	AR-1268-3	1.205					1.205	E6	0.00
44)	L9	AR-1268-4	5.044					5.044	E5	0.00
45)	L9	AR-1268-5	3.912					3.912	E6	0.00

Signal #2 Calibration Files

500 =PQ041998.D 1000 =PQ041996.D 750 =PQ041997.D
 250 =PQ041999.D 50 =PQ042000.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	2.358	2.757	2.387	2.744	2.616	2.573	E6	7.42
2) SA	Decachlorobiphenyl	6.668	6.550	6.671	7.048	7.701	6.928	E6	6.80
3) L1	AR-1016-1	1.261	1.298	1.183	1.457	1.622	1.364	E5	12.85
4) L1	AR-1016-2	1.827	1.882	1.751	2.109	2.128	1.939	E5	8.77
5) L1	AR-1016-3	0.991	1.016	0.939	1.128	1.146	1.044	E5	8.61

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ080519.M
 Title : GC EXTRACTABLES
 Last Update : Tue Aug 06 01:47:44 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ041998.D 1000 =PQ041996.D 750 =PQ041997.D
 250 =PQ041999.D 50 =PQ042000.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	7.906	7.884	7.424	9.249	8.915	8.276	E4	9.31
7)	L1 AR-1016-5	1.063	1.068	1.008	1.218	1.141	1.100	E5	7.43
8)	L2 AR-1221-1	2.887					2.887	E4	0.00
9)	L2 AR-1221-2	2.316					2.316	E4	0.00
10)	L2 AR-1221-3	7.453					7.453	E4	0.00
11)	L3 AR-1232-1	6.158					6.158	E4	0.00
12)	L3 AR-1232-2	7.835					7.835	E4	0.00
13)	L3 AR-1232-3	4.134					4.134	E4	0.00
14)	L3 AR-1232-4	3.627					3.627	E4	0.00
15)	L3 AR-1232-5	4.066					4.066	E4	0.00
16)	L4 AR-1242-1	1.041					1.041	E5	0.00
17)	L4 AR-1242-2	1.497					1.497	E5	0.00
18)	L4 AR-1242-3	8.074					8.074	E4	0.00
19)	L4 AR-1242-4	7.925					7.925	E4	0.00
20)	L4 AR-1242-5	1.234					1.234	E5	0.00
21)	L5 AR-1248-1	7.270					7.270	E4	0.00
22)	L5 AR-1248-2	9.986					9.986	E4	0.00
23)	L5 AR-1248-3	1.037					1.037	E5	0.00
24)	L5 AR-1248-4	1.277					1.277	E5	0.00
25)	L5 AR-1248-5	1.586					1.586	E5	0.00
26)	L6 AR-1254-1	2.379					2.379	E5	0.00
27)	L6 AR-1254-2	2.147					2.147	E5	0.00
28)	L6 AR-1254-3	3.935					3.935	E5	0.00
29)	L6 AR-1254-4	2.737					2.737	E5	0.00
30)	L6 AR-1254-5	3.877					3.877	E5	0.00
31)	L7 AR-1260-1	2.480	2.409	2.335	2.779	3.138	2.628	E5	12.61
32)	L7 AR-1260-2	3.353	3.284	3.189	3.668	4.126	3.524	E5	10.83
33)	L7 AR-1260-3	3.014	2.989	2.882	3.281	3.590	3.151	E5	9.07
34)	L7 AR-1260-4	2.827	2.689	2.725	3.046	3.216	2.901	E5	7.74
35)	L7 AR-1260-5	7.746	7.696	7.622	8.111	8.296	7.894	E5	3.72
36)	L8 AR-1262-1	3.665					3.665	E5	0.00
37)	L8 AR-1262-2	7.196					7.196	E5	0.00
38)	L8 AR-1262-3	3.061					3.061	E5	0.00
39)	L8 AR-1262-4	5.471					5.471	E5	0.00
40)	L8 AR-1262-5	2.633					2.633	E5	0.00
41)	L9 AR-1268-1	1.250					1.250	E6	0.00
42)	L9 AR-1268-2	1.150					1.150	E6	0.00
43)	L9 AR-1268-3	9.872					9.872	E5	0.00
44)	L9 AR-1268-4	4.019					4.019	E5	0.00
45)	L9 AR-1268-5	2.918					2.918	E6	0.00

(#) = Out of Range