

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080519\
 Data File : PQ042005.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Aug 2019 14:39
 Operator : SM\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 06 01:33:44 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080519.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 06 01:33:26 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.692	4.752	144.4E6	117.7E6	50.000	50.000
2)	SA Decachlor...	12.086	10.503	415.5E6	255.1E6	50.000	50.000
Target Compounds							
26)	L6 AR-1254-1	8.096	7.155	94291238	118.9E6	500.000	500.000
27)	L6 AR-1254-2	8.335	7.322	159.7E6	107.4E6	500.000	500.000
28)	L6 AR-1254-3	8.728	7.770	195.7E6	196.8E6	500.000	500.000
29)	L6 AR-1254-4	9.031	8.018	163.1E6	136.9E6	500.000	500.000
30)	L6 AR-1254-5	9.469	8.464	182.7E6	193.9E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080519\
Data File : PQ042005.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Aug 2019 14:39
Operator : SM\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 01:33:44 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080519.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 06 01:33:26 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

