

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080619\
Data File : PQ042015.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2019 12:49
Operator : SM\AJ
Sample : AIBLK37
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_Q
LabSampleId :
AIBLK37

Manual Integrations
APPROVED

Ankita
8/6/2019 2:39:13 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 13:34:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072719CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 29 04:03:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

5.682min 22.429 ng/ml

response 51439336

(1) Tetrachloro-m-xylene #2 (SA)

4.741min 25.269 ng/ml

response 48699897

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080619\
Data File : PQ042015.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2019 12:49
Operator : SM\AJ
Sample : AIBLK37
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
ECD_Q
LabSampleId :
AIBLK37

Manual Integrations
APPROVED

Ankita
8/6/2019 2:39:13 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 13:34:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072719CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 29 04:03:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(1) Tetrachloro-m-xylene (SA)

5.681min 24.742 ng/ml m

response 56742694

(1) Tetrachloro-m-xylene #2 (SA)

4.741min 27.142 ng/ml m

response 52309996

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080619\
Data File : PQ042015.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Aug 2019 12:49
Operator : SM\AJ
Sample : AIBLK37
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_Q
LabSampleId :
AIBLK37

Manual Integrations
APPROVED

Ankita
8/6/2019 2:39:13 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 06 13:34:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072719CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 29 04:03:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.681	4.741	56742694	52309996	24.742m>	27.142m>
2) SA Decachlor...	12.074	10.496	346.2E6	254.1E6	41.063	52.535 #

AJ
08/06/19

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.