

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080919\
Data File : PQ042200.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2019 22:03
Operator : SM\AJ
Sample : K4041-07 (Sig #1); K4041-07RX (Sig #2)
Misc :
ALS Vial : 54 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
BENT1

Manual Integrations
APPROVED

Ankita
8/12/2019 9:17:59 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 21 11:27:12 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072719CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 29 04:03:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
8.08 2898717 20.84
8.32 4955087 21.65
8.72 6278015 22.96
9.02 2763945 11.23
9.46 3771503 13.92

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
7.14 3636265 19.88
7.31 3763776 23.25
7.76 7388623 25.62
8.01 2043900 9.42
8.45 3232871 10.75

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080919\
Data File : PQ042200.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Aug 2019 22:03
Operator : SM\AJ
Sample : K4041-07 (Sig #1); K4041-07RX (Sig #2)
Misc :
ALS Vial : 54 Sample Multiplier: 1


Instrument :
ECD_Q
ClientSampleId :
BENT1

Manual Integrations
APPROVED

Ankita
8/12/2019 9:17:59 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 09 05:19:34 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072719CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jul 29 04:03:40 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
8.08 2488598 17.89
8.32 4955087 21.65
8.72 6278015 22.96
9.02 2382888 9.68
9.46 3274753 12.09
(26) AR-1254-1 #2 (L6)
R.T. Response Conc
7.14 4124646 22.55
7.31 3720223 22.98
7.76 7553616 26.19
8.01 2028230 9.34
8.45 3196862 10.63

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080919\
 Data File : PQ042200.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Aug 2019 22:03
 Operator : SM\AJ
 Sample : K4041-07
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 BENT1

Manual Integrations
 APPROVED

Ankita
 8/12/2019 9:17:59 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 09 05:19:34 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ072719CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jul 29 04:03:40 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.678	4.738	20052351	17216912	8.744	8.933
2) SA Decachlor...	12.065	10.487	61527205	48109584	7.297	9.945 #
Target Compounds						
26) L6 AR-1254-1	8.084	7.141	2488598	4124646	17.889	22.553 #
27) L6 AR-1254-2	8.324	7.309	4955087	3720223	21.649	22.983
28) L6 AR-1254-3	8.718	7.758	6278015	7553616	22.961	26.191
29) L6 AR-1254-4	9.018	8.006	2382888	2028230	9.680	9.344
30) L6 AR-1254-5	9.459	8.453	3274753	3196862	12.091m	10.629 >

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

AD
 08/21/19
