

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080922\
 Data File : PQ058877.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Aug 2022 00:26
 Operator : YP\AJ
 Sample : AR1254ICC800
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12544039

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/10/2022
 Supervised By :Ankita Jodhani 08/10/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 02:31:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080922CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 10 02:22:35 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.683	4.046	335.7E6	280.0E6	40.044	40.030
2) SA Decachlor...	10.586	9.160	548.7E6	321.8E6	78.702	78.320
Target Compounds						
26) L6 AR-1254-1	6.721	5.951	169.6E6	263.1E6	790.116	787.476
27) L6 AR-1254-2	6.938	6.096	258.0E6	225.7E6	787.952	785.879
28) L6 AR-1254-3	7.308	6.504	301.6E6	359.0E6	786.224	785.406
29) L6 AR-1254-4	7.595	6.730	182.4E6	196.0E6	784.459	795.695m
30) L6 AR-1254-5	8.014	7.150	217.3E6	283.5E6	782.248	782.786

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080922\
Data File : PQ058877.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Aug 2022 00:26
Operator : YP\AJ
Sample : AR1254ICC800
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12544039

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/10/2022
Supervised By :Ankita Jodhani 08/10/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 02:31:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080922CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 10 02:22:35 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

