

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080923\
 Data File : PQ062910.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Aug 2023 00:20
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 03:25:43 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 10 03:17:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.454	2.785	570.4E6	480.9E6	97.844	97.189
2) SA Decachlor...	8.687	7.595	704.7E6	779.1E6	96.653	96.845
Target Compounds						
41) L9 AR-1268-1	7.559	6.549	617.8E6	712.3E6	968.246	970.144
42) L9 AR-1268-2	7.637	6.612	567.6E6	659.0E6	966.740	972.477
43) L9 AR-1268-3	7.819	6.815	473.6E6	560.9E6	969.884	971.037
44) L9 AR-1268-4	8.143	7.106	205.6E6	241.4E6	962.236	961.232
45) L9 AR-1268-5	8.445	7.376	1520.2E6	1741.7E6	979.399	979.368

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080923\
Data File : PQ062910.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Aug 2023 00:20
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 03:25:43 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080923.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 10 03:17:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

