

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080923\
 Data File : PQ062907.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Aug 2023 23:36
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 10 03:04:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 10 02:53:49 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.454	2.786	146.8E6	124.8E6	25.683	25.891
2) SA Decachlor...	8.686	7.599	110.7E6	122.3E6	26.431	26.551
Target Compounds						
26) L6 AR-1254-1	5.396	4.563	58534371	71054911	269.045	267.937
27) L6 AR-1254-2	5.614	4.710	90699809	62179368	269.315	269.175
28) L6 AR-1254-3	5.971	5.096	94301961	99499453	265.947	264.368
29) L6 AR-1254-4	6.257	5.323	73340948	64826267	271.186	266.068
30) L6 AR-1254-5	6.672	5.733	73211516	91509877	264.901	265.035


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ080923\
Data File : PQ062907.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Aug 2023 23:36
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 10 03:04:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080923.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 10 02:53:49 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.454 min
Delta R.T.: 0.000 min
Response: 146781544
Conc: 25.68 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 2.786 min
Delta R.T.: 0.000 min
Response: 124750056
Conc: 25.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.686 min
Delta R.T.: 0.001 min
Response: 110730118
Conc: 26.43 ng/ml



#2 Decachlorobiphenyl


R.T.: 7.599 min
Delta R.T.: 0.000 min
Response: 122269568
Conc: 26.55 ng/ml



#26 AR-1254-1


R.T.: 5.396 min
Delta R.T.: 0.000 min
Response: 58534371
Conc: 269.04 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 4.563 min
Delta R.T.: 0.002 min
Response: 71054911
Conc: 267.94 ng/ml




#27 AR-1254-2

R.T.: 5.614 min
Delta R.T.: 0.001 min
Response: 90699809
Conc: 269.32 ng/ml



#27 AR-1254-2


R.T.: 4.710 min
Delta R.T.: 0.000 min
Response: 62179368
Conc: 269.18 ng/ml



#28 AR-1254-3


R.T.: 5.971 min
Delta R.T.: 0.002 min
Response: 94301961
Conc: 265.95 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 5.096 min
Delta R.T.: 0.001 min
Response: 99499453
Conc: 264.37 ng/ml




#29 AR-1254-4

R.T.: 6.257 min
Delta R.T.: 0.000 min
Response: 73340948
Conc: 271.19 ng/ml



#29 AR-1254-4


R.T.: 5.323 min
Delta R.T.: 0.002 min
Response: 64826267
Conc: 266.07 ng/ml



#30 AR-1254-5

R.T.: 6.672 min
Delta R.T.: 0.000 min
Response: 73211516
Conc: 264.90 ng/ml

Instrument :
ECD_Q
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 5.733 min
Delta R.T.: 0.000 min
Response: 91509877
Conc: 265.03 ng/ml