

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ081324\
 Data File : PQ067928.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Aug 2024 04:45
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/14/2024
 Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 13 10:18:13 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 09 02:37:44 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.459	2.825	305.1E6	336.5E6	51.421	53.061
2) SA Decachlor...	8.738	7.607	205.0E6	279.0E6	53.198	49.500
Target Compounds						
3) L1 AR-1016-1	4.573	3.833	91651164	109.6E6	554.492	622.089
4) L1 AR-1016-2	4.592	3.849	148.4E6	167.3E6	504.079	573.671
5) L1 AR-1016-3	4.651	4.010	88464464	93694862	463.467	595.166 #
6) L1 AR-1016-4	4.744	4.058	72122555	80960070	479.398	544.815
7) L1 AR-1016-5	5.035	4.255	68659656	97011795	469.743	567.104
31) L7 AR-1260-1	6.156	5.252	137.7E6	179.9E6	502.597	529.208
32) L7 AR-1260-2	6.422	5.442	169.5E6	212.2E6	525.644	542.078
33) L7 AR-1260-3	6.784	5.584	108.2E6	200.7E6	472.261	533.964
34) L7 AR-1260-4	7.003	6.046	123.2E6	150.8E6	482.932m	521.443
35) L7 AR-1260-5	7.323	6.292	238.4E6	323.6E6	538.602	543.221

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ081324\
Data File : PQ067928.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Aug 2024 04:45
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/14/2024
Supervised By :Ankita Jodhani 08/14/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 13 10:18:13 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080524.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 09 02:37:44 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

