

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ081523CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Aug 16 02:26:14 2023
 Response Via : Initial Calibration

Calibration Files

400 =PQ063071.D 1600 =PQ063073.D 800 =PQ063072.D
 200 =PQ063070.D 100 =PQ063069.D

Compound			400	1600	800	200	100	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	6.197	5.777	5.936	6.488	6.764	6.232	E6	6.45
2)	SA	Decachlorobiphenyl	4.799	4.237	4.491	5.247	5.604	4.876	E6	11.37
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1	2.484	2.155	2.270	2.701	2.876	2.497	E5	11.91
27)	L6	AR-1254-2	3.882	3.394	3.574	4.235	4.538	3.924	E5	11.95
28)	L6	AR-1254-3	4.110	3.686	3.842	4.432	4.689	4.152	E5	9.95
29)	L6	AR-1254-4	3.125	2.767	2.903	3.383	3.613	3.158	E5	10.94
30)	L6	AR-1254-5	3.371	3.031	3.117	3.590	3.782	3.378	E5	9.32
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1						0.000		-1.00
42)	L9	AR-1268-2						0.000		-1.00
43)	L9	AR-1268-3						0.000		-1.00
44)	L9	AR-1268-4						0.000		-1.00
45)	L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ063071.D 1600 =PQ063073.D 800 =PQ063072.D
 200 =PQ063070.D 100 =PQ063069.D

Compound		400	1600	800	200	100	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	4.682	4.148	4.371	5.124	5.512	4.767	E6	11.63
2) SA	Decachlorobiphenyl	4.824	4.205	4.446	5.250	5.643	4.874	E6	11.99
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ081523CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Aug 16 02:26:14 2023
 Response Via : Initial Calibration

Calibration Files

400 =PQ063071.D 1600 =PQ063073.D 800 =PQ063072.D
 200 =PQ063070.D 100 =PQ063069.D

	Compound	400	1600	800	200	100	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1	2.718	2.345	2.478	2.951	3.189	2.736 E5	12.55
27) L6	AR-1254-2	2.398	2.043	2.170	2.601	2.827	2.408 E5	13.18
28) L6	AR-1254-3	3.868	3.393	3.568	4.193	4.482	3.901 E5	11.42
29) L6	AR-1254-4	2.693	2.348	2.485	2.917	3.119	2.712 E5	11.54
30) L6	AR-1254-5	3.769	3.312	3.501	4.081	4.261	3.785 E5	10.39
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range