

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ081523CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Aug 16 02:51:10 2023
 Response Via : Initial Calibration

Calibration Files

400 =PQ063076.D 1600 =PQ063078.D 800 =PQ063077.D
 200 =PQ063075.D 100 =PQ063074.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	6.373	5.702	6.047	6.606	6.984	6.342	E6	7.80
2) SA	Decachlorobiphenyl	4.964	4.215	4.578	5.409	5.747	4.983	E6	12.37
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1	3.980	3.482	3.739	4.297	4.575	4.014	E5	10.82
37) L8	AR-1262-2	6.912	6.126	6.478	7.424	7.735	6.935	E5	9.51
38) L8	AR-1262-3	4.606	3.995	4.262	4.987	5.207	4.612	E5	10.83
39) L8	AR-1262-4	3.565	3.018	3.266	3.843	3.988	3.536	E5	11.32
40) L8	AR-1262-5	2.362	2.031	2.201	2.576	2.702	2.374	E5	11.45
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

400 =PQ063076.D 1600 =PQ063078.D 800 =PQ063077.D
 200 =PQ063075.D 100 =PQ063074.D

	Compound	400	1600	800	200	100	Avg		%RSD
1) SA	Tetrachloro-m-xylene	4.773	4.055	4.381	5.247	5.632	4.818	E6	13.23
2) SA	Decachlorobiphenyl	4.889	4.191	4.523	5.384	5.723	4.942	E6	12.59
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ081523CLP.M
 Title : GC EXTRACTABLES
 Last Update : Wed Aug 16 02:51:10 2023
 Response Via : Initial Calibration

Calibration Files

400 =PQ063076.D 1600 =PQ063078.D 800 =PQ063077.D
 200 =PQ063075.D 100 =PQ063074.D

Compound			400	1600	800	200	100	Avg	%RSD

6) L1	AR-1016-4							0.000	-1.00
7) L1	AR-1016-5							0.000	-1.00
8) L2	AR-1221-1							0.000	-1.00
9) L2	AR-1221-2							0.000	-1.00
10) L2	AR-1221-3							0.000	-1.00
11) L3	AR-1232-1							0.000	-1.00
12) L3	AR-1232-2							0.000	-1.00
13) L3	AR-1232-3							0.000	-1.00
14) L3	AR-1232-4							0.000	-1.00
15) L3	AR-1232-5							0.000	-1.00
16) L4	AR-1242-1							0.000	-1.00
17) L4	AR-1242-2							0.000	-1.00
18) L4	AR-1242-3							0.000	-1.00
19) L4	AR-1242-4							0.000	-1.00
20) L4	AR-1242-5							0.000	-1.00
21) L5	AR-1248-1							0.000	-1.00
22) L5	AR-1248-2							0.000	-1.00
23) L5	AR-1248-3							0.000	-1.00
24) L5	AR-1248-4							0.000	-1.00
25) L5	AR-1248-5							0.000	-1.00
26) L6	AR-1254-1							0.000	-1.00
27) L6	AR-1254-2							0.000	-1.00
28) L6	AR-1254-3							0.000	-1.00
29) L6	AR-1254-4							0.000	-1.00
30) L6	AR-1254-5							0.000	-1.00
31) L7	AR-1260-1							0.000	-1.00
32) L7	AR-1260-2							0.000	-1.00
33) L7	AR-1260-3							0.000	-1.00
34) L7	AR-1260-4							0.000	-1.00
35) L7	AR-1260-5							0.000	-1.00
36) L8	AR-1262-1		3.776	3.214	3.464	4.143	4.330	3.786 E5	12.20
37) L8	AR-1262-2		6.760	6.016	6.364	7.247	7.399	6.757 E5	8.61
38) L8	AR-1262-3		2.764	2.352	2.538	3.050	3.127	2.766 E5	11.91
39) L8	AR-1262-4		5.069	4.459	4.760	5.486	5.697	5.094 E5	9.98
40) L8	AR-1262-5		2.460	2.122	2.278	2.726	2.859	2.489 E5	12.27
41) L9	AR-1268-1							0.000	-1.00
42) L9	AR-1268-2							0.000	-1.00
43) L9	AR-1268-3							0.000	-1.00
44) L9	AR-1268-4							0.000	-1.00
45) L9	AR-1268-5							0.000	-1.00

 (#) = Out of Range