

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ081523\
 Data File : PQ063063.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2023 17:59
 Operator : YP\AJ
 Sample : AR1242ICC1600
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12425028

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 15 18:14:35 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ081523CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Aug 15 17:58:54 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.454	2.784	462.1E6	326.8E6	74.615	70.097
2) SA Decachlor...	8.686	7.593	660.8E6	663.2E6	138.206	138.649
Target Compounds						
16) L4 AR-1242-1	4.566	3.795	252.4E6	188.5E6	1389.769	1366.663
17) L4 AR-1242-2	4.586	3.812	363.8E6	255.9E6	1412.386	1374.691
18) L4 AR-1242-3	4.644	3.973	223.6E6	148.5E6	1360.548	1373.022
19) L4 AR-1242-4	4.737	4.059	188.3E6	133.1E6	1398.242	1348.342
20) L4 AR-1242-5	5.458	4.559	204.7E6	182.1E6	1402.828	1362.493

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ081523\
Data File : PQ063063.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2023 17:59
Operator : YP\AJ
Sample : AR1242ICC1600
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12425028

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 15 18:14:35 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ081523CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Aug 15 17:58:54 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

