

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ081523\
 Data File : PQ063083.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2023 22:52
 Operator : YP\AJ
 Sample : AR1268ICC1600
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12685048

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 03:01:20 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ081523CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Aug 16 02:55:24 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.453	2.783	451.5E6	321.3E6	77.221	73.535
2) SA Decachlor...	8.685	7.592	1181.0E6	1181.0E6	148.615	149.378
Target Compounds						
41) L9 AR-1268-1	7.558	6.547	1092.9E6	1098.1E6	1491.230	1486.870
42) L9 AR-1268-2	7.636	6.610	1001.4E6	1001.9E6	1494.541	1501.829
43) L9 AR-1268-3	7.819	6.812	831.1E6	856.0E6	1497.562	1495.523
44) L9 AR-1268-4	8.144	7.104	360.0E6	376.6E6	1467.890	1473.753
45) L9 AR-1268-5	8.445	7.373	2634.5E6	2693.5E6	1528.641	1521.154

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ081523\
Data File : PQ063083.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2023 22:52
Operator : YP\AJ
Sample : AR1268ICC1600
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12685048

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 03:01:20 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ081523CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Aug 16 02:55:24 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

