

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ081524\
 Data File : PQ067976.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Aug 2024 11:16
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/15/2024
 Supervised By :Ankita Jodhani 08/15/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 15 04:50:23 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 09 02:37:44 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.461	2.825	275.0E6	294.6E6	46.350	46.447
2) SA Decachlor...	8.753	7.607	188.5E6	266.6E6	48.911	47.304
Target Compounds						
3) L1 AR-1016-1	4.577	3.832	79905020	91594606	483.427	519.669
4) L1 AR-1016-2	4.596	3.848	134.7E6	146.3E6	457.376	501.657
5) L1 AR-1016-3	4.655	4.010	82418370	79313042	431.792	503.810
6) L1 AR-1016-4	4.748	4.058	65831496	70367736	437.581	473.534
7) L1 AR-1016-5	5.040	4.254	68151598	82928514	466.267	484.777
31) L7 AR-1260-1	6.163	5.252	128.2E6	158.1E6	467.928	465.222
32) L7 AR-1260-2	6.429	5.441	151.0E6	185.5E6	468.369m	473.748
33) L7 AR-1260-3	6.790	5.583	100.7E6	179.9E6	439.888	478.830
34) L7 AR-1260-4	7.011	6.046	117.4E6	136.2E6	460.184m	470.794
35) L7 AR-1260-5	7.333	6.291	217.4E6	301.5E6	491.181	506.062

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ081524\
Data File : PQ067976.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Aug 2024 11:16
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/15/2024
Supervised By :Ankita Jodhani 08/15/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 15 04:50:23 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080524.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 09 02:37:44 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

