

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ081618\
Data File : PQ031226.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2018 09:35
Operator : SM\SJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_Q
LabSampleId :
AR1254310

Manual Integrations
APPROVED

Sohil
8/17/2018 11:25:45 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 11:28:35 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080918CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 13 07:57:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



(26) AR-1254-1 (L6)
R.T. Response Conc
6.86 62396119 390.25
7.14 38788236 394.24
7.23 71214717 398.29
7.51 52248212 399.71
7.79 39777206 396.65

(26) AR-1254-1 #2 (L6)
R.T. Response Conc
5.94 44961518 317.33
6.26 38886278 325.19
6.58 60094308 338.00
6.89 36063093 336.42
6.99 73895755 334.95

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ081618\
Data File : PQ031226.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Aug 2018 09:35
Operator : SM\SJ
Sample : AR1254CCC400
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
ECD_Q
LabSampleId :
AR1254310

Manual Integrations
APPROVED

Sohil
8/17/2018 11:25:45 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 11:28:35 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080918CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 13 07:57:53 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm



Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ081618\
 Data File : PQ031226.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Aug 2018 09:35
 Operator : SM\SJ
 Sample : AR1254CCC400
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_Q
 LabSampleId :
 AR1254310

Manual Integrations
 APPROVED

Sohil
 8/17/2018 11:25:45 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 11:28:35 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080918CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 13 07:57:53 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.605	3.886	35171240	31081694	17.692	15.995
2) SA Decachlor...	10.460	8.972	103.3E6	101.3E6	51.929	45.941
Target Compounds						
26) L6 AR-1254-1	6.856	5.940	62396119	44961518	390.251	317.335
27) L6 AR-1254-2	7.140	6.258	38788236	38886278	394.237	325.192
28) L6 AR-1254-3	7.225	6.575	71214717	54113484	398.295	304.357m
29) L6 AR-1254-4	7.510	6.894	52248212	36063093	399.712	336.424
30) L6 AR-1254-5	7.786	6.995	39777206	73895755	396.649	334.951

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

> SJ 08/17/18