

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ081624\
 Data File : PQ068021.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Aug 2024 18:24
 Operator : YP\AJ
 Sample : P3610-01
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 A6934

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 16 04:54:09 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080524.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 09 02:37:44 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.458	2.825	120.0E6	129.1E6	20.227	20.353
2) SA Decachlor...	8.735	7.606	57945387	84805804	15.039	15.047

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ081624\
Data File : PQ068021.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Aug 2024 18:24
Operator : YP\AJ
Sample : P3610-01
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
A6934

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 16 04:54:09 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ080524.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 09 02:37:44 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.458 min
Delta R.T.: -0.003 min
Response: 120009003
Conc: 20.23 ng/ml

Instrument :
ECD_Q
ClientSampleId :
A6934




#1 Tetrachloro-m-xylene

R.T.: 2.825 min
Delta R.T.: -0.004 min
Response: 129081148
Conc: 20.35 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.735 min
Delta R.T.: -0.001 min
Response: 57945387
Conc: 15.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 7.606 min
Delta R.T.: -0.006 min
Response: 84805804
Conc: 15.05 ng/ml