

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082124\
 Data File : PQ068074.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2024 16:54
 Operator : YP\AJ
 Sample : AR1232ICC1600
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12325013

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 03:30:37 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 22 03:25:44 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.439	2.823	538.1E6	543.0E6	72.898	73.830
2) SA Decachlor...	8.612	7.604	306.9E6	487.4E6	149.599	137.406
Target Compounds						
11) L3 AR-1232-1	3.787	3.169	194.9E6	204.5E6	1402.529	1345.862
12) L3 AR-1232-2	4.281	3.845	102.4E6	221.8E6	1422.590	1404.155
13) L3 AR-1232-3	4.558	4.006	207.0E6	113.2E6	1411.463	1403.813
14) L3 AR-1232-4	4.708	4.093	102.3E6	94986278	1404.320	1371.392
15) L3 AR-1232-5	4.805	4.251	64622027	107.8E6	1426.263	1342.799

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082124\
Data File : PQ068074.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2024 16:54
Operator : YP\AJ
Sample : AR1232ICC1600
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12325013

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 03:30:37 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 22 03:25:44 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

