

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082124\
 Data File : PQ068078.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2024 17:52
 Operator : YP\AJ
 Sample : AR1242ICC800
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12424017

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 03:45:03 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 22 03:40:00 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.439	2.823	279.3E6	282.3E6	37.121	37.117
2) SA Decachlor...	8.612	7.604	162.2E6	266.8E6	76.242	73.652
Target Compounds						
16) L4 AR-1242-1	4.540	3.830	134.2E6	145.2E6	744.368	716.683
17) L4 AR-1242-2	4.559	3.846	207.2E6	219.7E6	721.919	730.570
18) L4 AR-1242-3	4.617	4.007	125.7E6	115.7E6	724.933	728.374
19) L4 AR-1242-4	4.709	4.093	105.0E6	108.4E6	735.221	720.910
20) L4 AR-1242-5	5.423	4.590	109.3E6	144.3E6	738.221	726.415

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082124\
Data File : PQ068078.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2024 17:52
Operator : YP\AJ
Sample : AR1242ICC800
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12424017

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 03:45:03 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 22 03:40:00 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

