

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082124\
 Data File : PQ068084.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2024 19:20
 Operator : YP\AJ
 Sample : AR1248ICC1600
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12485023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 04:05:31 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 22 03:56:19 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.439	2.823	524.0E6	545.7E6	70.789	71.817
2) SA Decachlor...	8.612	7.605	310.8E6	487.7E6	146.755	137.393
Target Compounds						
21) L5 AR-1248-1	4.539	3.829	189.9E6	197.7E6	1377.508	1360.474
22) L5 AR-1248-2	4.804	4.055	251.2E6	298.2E6	1366.304	1348.803
23) L5 AR-1248-3	4.995	4.093	308.7E6	289.3E6	1377.806	1344.993
24) L5 AR-1248-4	5.388	4.251	349.2E6	361.4E6	1405.879	1362.265
25) L5 AR-1248-5	5.423	4.628	348.6E6	367.0E6	1400.605	1375.917

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082124\
Data File : PQ068084.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2024 19:20
Operator : YP\AJ
Sample : AR1248ICC1600
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12485023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 04:05:31 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 22 03:56:19 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

