

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082124\
 Data File : PQ068083.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Aug 2024 19:05
 Operator : YP\AJ
 Sample : AR1248ICC800
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12484022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 22 04:05:06 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 22 03:56:19 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.439	2.823	278.2E6	289.5E6	37.585	38.095
2) SA Decachlor...	8.611	7.604	165.7E6	268.2E6	78.231	75.575
Target Compounds						
21) L5 AR-1248-1	4.539	3.830	101.3E6	112.0E6	734.816	770.688
22) L5 AR-1248-2	4.805	4.056	136.0E6	164.1E6	739.986	741.963
23) L5 AR-1248-3	4.995	4.093	166.5E6	159.1E6	743.206	739.430
24) L5 AR-1248-4	5.389	4.251	187.7E6	198.3E6	755.841	747.659
25) L5 AR-1248-5	5.424	4.628	187.3E6	199.8E6	752.675	748.904

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082124\
Data File : PQ068083.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Aug 2024 19:05
Operator : YP\AJ
Sample : AR1248ICC800
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12484022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 22 04:05:06 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 22 03:56:19 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

