

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082324\
 Data File : PQ068160.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 23 Aug 2024 13:55
 Operator : YP\AJ
 Sample : P3673-22MSD
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 DCXZ4MSD

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/26/2024
 Supervised By :Ankita Jodhani 08/26/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 24 05:29:41 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Aug 22 09:07:54 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.438	2.822	116.5E6	106.9E6	18.538	15.903
2) SA Decachlor...	8.609	7.601	60285727	95261569	35.430	30.570
Target Compounds						
3) L1 AR-1016-1	4.538	3.828	66898074	69199406	372.902	331.879
4) L1 AR-1016-2	4.558	3.844	110.9E6	102.9E6	396.238	328.865
5) L1 AR-1016-3	4.615	4.005	75315266	53521968	443.882	324.918 #
6) L1 AR-1016-4	4.707	4.054	58582387	44961989	417.731	333.621
7) L1 AR-1016-5	4.994	4.250	48378172	60155772	381.624	346.032
31) L7 AR-1260-1	6.095	5.246	94107497	98030958	364.680	283.412m
32) L7 AR-1260-2	6.353	5.437	101.4E6	93634257	361.509	225.071m#
33) L7 AR-1260-3	6.706	5.579	63837524	110.5E6	299.310	272.103
34) L7 AR-1260-4	6.924	6.040	73152250	79709038	314.287	247.678m
35) L7 AR-1260-5	7.234	6.286	129.0E6	166.5E6	285.972	225.579

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082324\
Data File : PQ068160.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 23 Aug 2024 13:55
Operator : YP\AJ
Sample : P3673-22MSD
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :

ECD_Q

ClientSampleId :

DCXZ4MSD

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/26/2024

Supervised By :Ankita Jodhani 08/26/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 24 05:29:41 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Aug 22 09:07:54 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

