

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082819\
 Data File : PQ043056.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Aug 2019 20:42
 Operator : SM\AJ
 Sample : K4556-09
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 C0AR1

Manual Integrations
 APPROVED

Ankita
 8/28/2019 4:13:36 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 28 08:28:37 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082619CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Aug 26 05:38:16 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	5.622	4.679	1653.8E6	128.7E6	617.885m	55.611m#
2) SA Decachlor...	11.972	10.427	1343.9E6	774.4E6	159.190	115.274m#
Target Compounds						
26) L6 AR-1254-1	8.036	7.091	42486.1E6	40444.2E6	254254.270	164909.850 #
27) L6 AR-1254-2	8.280	7.263	55455.8E6	42588.9E6	206192.750	190947.079
28) L6 AR-1254-3	8.670	7.710	51025.8E6	57599.7E6	169420.526	153242.098
29) L6 AR-1254-4	8.971	7.956	49007.4E6	47767.9E6	215113.160	197552.765
30) L6 AR-1254-5	9.385	8.408	53495.8E6	69894.8E6	197528.765m	189390.458

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ082819\
Data File : PQ043056.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Aug 2019 20:42
Operator : SM\AJ
Sample : K4556-09
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampled :
C0AR1

Manual Integrations
APPROVED

Ankita
8/28/2019 4:13:36 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 28 08:28:37 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ082619CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Aug 26 05:38:16 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

