

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ083019\
 Data File : PQ043160.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Aug 2019 06:51
 Operator : SM\AJ
 Sample : AR1268ICC1600
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1268501

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Aug 30 07:06:23 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ083019CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Aug 30 06:59:33 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.710	4.806	258.9E6	140.6E6	79.661	77.049
2)	SA Decachlor...	12.102	10.596	1836.8E6	1174.6E6	154.954	154.828
Target Compounds							
41)	L9 AR-1268-1	10.482	9.390	1475.4E6	989.4E6	1571.398	1567.113
42)	L9 AR-1268-2	10.584	9.456	1368.1E6	931.6E6	1580.298	1584.620
43)	L9 AR-1268-3	10.830	9.676	1144.0E6	759.1E6	1585.310	1585.524
44)	L9 AR-1268-4	11.290	9.984	524.7E6	336.4E6	1540.152	1512.144
45)	L9 AR-1268-5	11.737	10.308	4141.2E6	2653.5E6	1585.876	1599.412

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ083019\
Data File : PQ043160.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Aug 2019 06:51
Operator : SM\AJ
Sample : AR1268ICC1600
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1268501

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Aug 30 07:06:23 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ083019CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Aug 30 06:59:33 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

