

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ090118.M
 Title : GC EXTRACTABLES
 Last Update : Fri Aug 31 04:27:04 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ031544.D 1000 =PQ031542.D 750 =PQ031543.D
 250 =PQ031545.D 50 =PQ031546.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.413	1.607	1.402	1.430	1.444	1.459	E6	5.77
2) SA	Decachlorobiphenyl	2.165	2.214	2.100	2.203	2.041	2.144	E6	3.41
3) L1	AR-1016-1	3.688	3.942	3.609	3.919	4.368	3.905	E4	7.58
4) L1	AR-1016-2	3.856	4.019	3.747	4.001	4.432	4.011	E4	6.49
5) L1	AR-1016-3	5.703	5.970	5.443	6.061	7.542	6.144	E4	13.32
6) L1	AR-1016-4	5.159	5.429	5.014	5.530	5.847	5.396	E4	6.03
7) L1	AR-1016-5	4.403	4.661	4.317	4.795	4.552	4.546	E4	4.23
8) L2	AR-1221-1	1.750					1.750	E4	0.00
9) L2	AR-1221-2	1.307					1.307	E4	0.00
10) L2	AR-1221-3	4.295					4.295	E4	0.00
11) L3	AR-1232-1	3.026					3.026	E4	0.00
12) L3	AR-1232-2	1.550					1.550	E4	0.00
13) L3	AR-1232-3	4.095					4.095	E3	0.00
14) L3	AR-1232-4	4.441					4.441	E3	0.00
15) L3	AR-1232-5	2.659					2.659	E3	0.00
16) L4	AR-1242-1	1.828					1.828	E4	0.00
17) L4	AR-1242-2	3.078					3.078	E4	0.00
18) L4	AR-1242-3	1.612					1.612	E4	0.00
19) L4	AR-1242-4	1.358					1.358	E4	0.00
20) L4	AR-1242-5	3.921					3.921	E4	0.00
21) L5	AR-1248-1	5.672					5.672	E4	0.00
22) L5	AR-1248-2	5.940					5.940	E4	0.00
23) L5	AR-1248-3	7.967					7.967	E4	0.00
24) L5	AR-1248-4	7.611					7.611	E4	0.00
25) L5	AR-1248-5	5.116					5.116	E4	0.00
26) L6	AR-1254-1	1.451					1.451	E5	0.00
27) L6	AR-1254-2	8.992					8.992	E4	0.00
28) L6	AR-1254-3	1.639					1.639	E5	0.00
29) L6	AR-1254-4	1.223					1.223	E5	0.00
30) L6	AR-1254-5	9.489					9.489	E4	0.00
31) L7	AR-1260-1	1.026	1.090	0.997	1.112	1.269	1.099	E5	9.63
32) L7	AR-1260-2	1.261	1.341	1.210	1.344	1.468	1.325	E5	7.38
33) L7	AR-1260-3	1.535	1.640	1.467	1.594	1.655	1.578	E5	4.94
34) L7	AR-1260-4	1.128	1.193	1.084	1.186	1.255	1.169	E5	5.58
35) L7	AR-1260-5	2.310	2.518	2.261	2.384	2.403	2.375	E5	4.14
36) L8	AR-1262-1	5.760					5.760	E4	0.00
37) L8	AR-1262-2	5.346					5.346	E4	0.00
38) L8	AR-1262-3	6.117					6.117	E4	0.00
39) L8	AR-1262-4	1.280					1.280	E5	0.00
40) L8	AR-1262-5	9.641					9.641	E4	0.00
41) L9	AR-1268-1	5.308					5.308	E4	0.00
42) L9	AR-1268-2	6.979					6.979	E4	0.00
43) L9	AR-1268-3	3.259					3.259	E5	0.00
44) L9	AR-1268-4	2.501					2.501	E5	0.00
45) L9	AR-1268-5	8.551					8.551	E5	0.00

Signal #2 Calibration Files

500 =PQ031544.D 1000 =PQ031542.D 750 =PQ031543.D
 250 =PQ031545.D 50 =PQ031546.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.085	1.265	1.064	1.089	1.195	1.139	E6	7.60
2) SA	Decachlorobiphenyl	2.064	2.137	2.001	2.126	2.150	2.096	E6	2.97
3) L1	AR-1016-1	3.042	3.381	3.003	3.164	3.451	3.208	E4	6.24
4) L1	AR-1016-2	4.723	5.234	4.693	4.941	5.482	5.015	E4	6.77
5) L1	AR-1016-3	7.052	7.541	6.815	7.214	7.412	7.207	E4	3.99

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ090118.M
 Title : GC EXTRACTABLES
 Last Update : Fri Aug 31 04:27:04 2018
 Response Via : Initial Calibration

Calibration Files

500 =PQ031544.D 1000 =PQ031542.D 750 =PQ031543.D
 250 =PQ031545.D 50 =PQ031546.D

	Compound	500	1000	750	250	50	Avg	%RSD
6)	L1 AR-1016-4	4.561	4.936	4.455	4.855	3.428	4.447 E4	13.58
7)	L1 AR-1016-5	3.907	4.280	3.902	4.085	3.878	4.011 E4	4.29
8)	L2 AR-1221-1	1.383					1.383 E4	0.00
9)	L2 AR-1221-2	1.089					1.089 E4	0.00
10)	L2 AR-1221-3	3.471					3.471 E4	0.00
11)	L3 AR-1232-1	1.261					1.261 E4	0.00
12)	L3 AR-1232-2	3.585					3.585 E3	0.00
13)	L3 AR-1232-3	1.233					1.233 E4	0.00
14)	L3 AR-1232-4	1.510					1.510 E4	0.00
15)	L3 AR-1232-5	2.599					2.599 E3	0.00
16)	L4 AR-1242-1	1.399					1.399 E4	0.00
17)	L4 AR-1242-2	2.870					2.870 E4	0.00
18)	L4 AR-1242-3	9.609					9.609 E3	0.00
19)	L4 AR-1242-4	8.795					8.795 E3	0.00
20)	L4 AR-1242-5	2.287					2.287 E4	0.00
21)	L5 AR-1248-1	4.721					4.721 E4	0.00
22)	L5 AR-1248-2	9.321					9.321 E4	0.00
23)	L5 AR-1248-3	6.392					6.392 E4	0.00
24)	L5 AR-1248-4	5.755					5.755 E4	0.00
25)	L5 AR-1248-5	4.618					4.618 E4	0.00
26)	L6 AR-1254-1	9.982					9.982 E4	0.00
27)	L6 AR-1254-2	8.769					8.769 E4	0.00
28)	L6 AR-1254-3	1.155					1.155 E5	0.00
29)	L6 AR-1254-4	8.304					8.304 E4	0.00
30)	L6 AR-1254-5	1.663					1.663 E5	0.00
31)	L7 AR-1260-1	0.971	1.059	0.949	1.044	1.027	1.010 E5	4.73
32)	L7 AR-1260-2	1.223	1.332	1.200	1.309	1.396	1.292 E5	6.23
33)	L7 AR-1260-3	1.163	1.274	1.137	1.234	1.206	1.202 E5	4.53
34)	L7 AR-1260-4	1.002	1.076	0.966	1.040	1.087	1.034 E5	4.89
35)	L7 AR-1260-5	2.509	2.744	2.462	2.557	2.461	2.546 E5	4.62
36)	L8 AR-1262-1	5.018					5.018 E4	0.00
37)	L8 AR-1262-2	5.442					5.442 E4	0.00
38)	L8 AR-1262-3	6.202					6.202 E4	0.00
39)	L8 AR-1262-4	1.047					1.047 E5	0.00
40)	L8 AR-1262-5	9.461					9.461 E4	0.00
41)	L9 AR-1268-1	5.385					5.385 E4	0.00
42)	L9 AR-1268-2	3.023					3.023 E5	0.00
43)	L9 AR-1268-3	2.502					2.502 E5	0.00
44)	L9 AR-1268-4	6.728					6.728 E4	0.00
45)	L9 AR-1268-5	8.421					8.421 E5	0.00

(#) = Out of Range