

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ090624\
 Data File : PQ068426.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Sep 2024 06:47
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 09/09/2024
 Supervised By :Ankita Jodhani 09/09/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 07 08:07:03 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ090624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 07 07:27:42 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.438	2.822	371.3E6	379.2E6	52.000	53.242
2)	SA Decachlor...	8.610	7.600	225.8E6	352.2E6	51.802	53.332
Target Compounds							
3)	L1 AR-1016-1	4.538	3.827	121.3E6	128.8E6	518.941	542.586
4)	L1 AR-1016-2	4.558	3.844	189.5E6	191.2E6	516.246	529.136
5)	L1 AR-1016-3	4.616	4.004	121.7E6	103.3E6	538.276	531.736
6)	L1 AR-1016-4	4.707	4.053	107.0E6	85893801	562.194	531.860
7)	L1 AR-1016-5	4.994	4.249	94236147	104.7E6	532.990m	541.982
31)	L7 AR-1260-1	6.095	5.246	171.8E6	204.7E6	516.520m	533.180m
32)	L7 AR-1260-2	6.354	5.436	193.1E6	249.4E6	515.182m	547.247
33)	L7 AR-1260-3	6.707	5.578	124.0E6	225.3E6	527.596	517.055
34)	L7 AR-1260-4	6.925	6.041	141.3E6	174.1E6	509.990	541.988
35)	L7 AR-1260-5	7.234	6.286	272.3E6	391.9E6	509.541	537.850

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ090624\
Data File : PQ068426.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Sep 2024 06:47
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR1660CCC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 09/09/2024
Supervised By :Ankita Jodhani 09/09/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 07 08:07:03 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ090624.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 07 07:27:42 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

