

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ090722\
 Data File : PQ058964.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Sep 2022 19:05
 Operator : YP\AJ
 Sample : AR1221ICC1600
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 08 01:16:06 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ090722CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 08 01:14:25 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
1)	SA Tetrachlo...	4.680	4.052	559.9E6	580.9E6	70.669	74.885
2)	SA Decachlor...	10.568	9.169	560.0E6	572.2E6	138.723	145.287
Target Compounds							
8)	L2 AR-1221-1	4.891	4.265	92283246	93579540	1356.357	1407.944
9)	L2 AR-1221-2	4.980	4.353	63991073	65280342	1335.002	1379.368
10)	L2 AR-1221-3	5.058	4.430	194.3E6	211.6E6	1308.843	1376.741


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ090722\
Data File : PQ058964.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Sep 2022 19:05
Operator : YP\AJ
Sample : AR1221ICC1600
Misc :
ALS Vial : 12 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 08 01:16:06 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ090722CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 08 01:14:25 2022
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.680 min
Delta R.T.: 0.000 min
Response: 559947879
Conc: 70.67 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 4.052 min
Delta R.T.: 0.000 min
Response: 580941882
Conc: 74.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.568 min
Delta R.T.: 0.001 min
Response: 559996913
Conc: 138.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.169 min
Delta R.T.: 0.001 min
Response: 572155686
Conc: 145.29 ng/ml




#8 AR-1221-1

R.T.: 4.891 min
Delta R.T.: 0.000 min
Response: 92283246
Conc: 1356.36 ng/ml




#8 AR-1221-1

R.T.: 4.265 min
Delta R.T.: 0.000 min
Response: 93579540
Conc: 1407.94 ng/ml




#9 AR-1221-2

R.T.: 4.980 min
Delta R.T.: 0.000 min
Response: 63991073
Conc: 1335.00 ng/ml




#9 AR-1221-2

R.T.: 4.353 min
Delta R.T.: 0.000 min
Response: 65280342
Conc: 1379.37 ng/ml



#10 AR-1221-3

R.T.: 5.058 min
Delta R.T.: 0.000 min
Response: 194337093
Conc: 1308.84 ng/ml



#10 AR-1221-3

R.T.: 4.430 min
Delta R.T.: 0.000 min
Response: 211580067
Conc: 1376.74 ng/ml