

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ090722\
 Data File : PQ058989.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 Sep 2022 01:09
 Operator : YP\AJ
 Sample : AR1262ICC1600
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 08 02:36:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ090722CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Sep 08 02:34:19 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds						
1) SA Tetrachlo...	4.680	4.051	468.2E6	568.5E6	69.328	74.968
2) SA Decachlor...	10.568	9.166	520.1E6	607.2E6	136.031	146.325
Target Compounds						
36) L8 AR-1262-1	8.077	7.153	414.4E6	265.4E6	1347.734	1471.862
37) L8 AR-1262-2	8.654	7.689	663.7E6	937.5E6	1425.031	1524.271
38) L8 AR-1262-3	8.978	7.974	308.7E6	357.4E6	1359.125	1469.462
39) L8 AR-1262-4	9.075	8.039	348.0E6	665.7E6	2530.748	1506.749 #
40) L8 AR-1262-5	9.776	8.562	218.8E6	276.8E6	1369.205	1485.077


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ090722\
Data File : PQ058989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 Sep 2022 01:09
Operator : YP\AJ
Sample : AR1262ICC1600
Misc :
ALS Vial : 37 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 08 02:36:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ090722CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Sep 08 02:34:19 2022
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.680 min
Delta R.T.: 0.000 min
Response: 468231549
Conc: 69.33 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 4.051 min
Delta R.T.: 0.000 min
Response: 568472352
Conc: 74.97 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.568 min
Delta R.T.: 0.000 min
Response: 520130668
Conc: 136.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.166 min
Delta R.T.: 0.000 min
Response: 607223248
Conc: 146.32 ng/ml




#36 AR-1262-1

R.T.: 8.077 min
Delta R.T.: 0.000 min
Response: 414427520
Conc: 1347.73 ng/ml




#36 AR-1262-1

R.T.: 7.153 min
Delta R.T.: 0.000 min
Response: 265402268
Conc: 1471.86 ng/ml




#37 AR-1262-2

R.T.: 8.654 min
Delta R.T.: 0.000 min
Response: 663662732
Conc: 1425.03 ng/ml




#37 AR-1262-2

R.T.: 7.689 min
Delta R.T.: 0.000 min
Response: 937480803
Conc: 1524.27 ng/ml




#38 AR-1262-3

R.T.: 8.978 min
Delta R.T.: 0.000 min
Response: 308667479
Conc: 1359.12 ng/ml




#38 AR-1262-3

R.T.: 7.974 min
Delta R.T.: 0.000 min
Response: 357407246
Conc: 1469.46 ng/ml




#39 AR-1262-4

R.T.: 9.075 min
Delta R.T.: 0.001 min
Response: 348028518
Conc: 2530.75 ng/ml




#39 AR-1262-4

R.T.: 8.039 min
Delta R.T.: 0.000 min
Response: 665743382
Conc: 1506.75 ng/ml



#40 AR-1262-5

R.T.: 9.776 min
Delta R.T.: 0.000 min
Response: 218771607
Conc: 1369.21 ng/ml



#40 AR-1262-5

R.T.: 8.562 min
Delta R.T.: 0.000 min
Response: 276841913
Conc: 1485.08 ng/ml