

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ090922\
 Data File : PQ059065.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Sep 2022 23:14
 Operator : YP\AJ
 Sample : AR1268ICC1600
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_Q
 ClientSampleId :
 AR12685020

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 10 15:13:57 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ090922CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Sep 10 15:11:41 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.696	4.050	811.3E6	523.4E6	72.161	75.769
2) SA Decachlor...	10.593	9.160	1323.2E6	1251.4E6	141.868	146.819
Target Compounds						
41) L9 AR-1268-1	8.997	7.970	1182.9E6	1197.4E6	1442.122	1504.949
42) L9 AR-1268-2	9.099	8.035	1046.0E6	1098.1E6	1446.149	1514.386
43) L9 AR-1268-3	9.337	8.250	945.2E6	949.6E6	1436.354	1512.721
44) L9 AR-1268-4	9.800	8.558	340.7E6	342.6E6	1409.363	1488.167
45) L9 AR-1268-5	10.237	8.878	2978.0E6	2960.1E6	1475.066	1523.425

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Data\PQ090922\
Data File : PQ059065.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Sep 2022 23:14
Operator : YP\AJ
Sample : AR1268ICC1600
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
ECD_Q
ClientSampleId :
AR12685020

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 10 15:13:57 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\PQ090922CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Sep 10 15:11:41 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

