

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ091119.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 12 02:33:14 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ043455.D 1000 =PQ043453.D 750 =PQ043454.D
 250 =PQ043456.D 50 =PQ043457.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	1.491	1.593	1.411	1.763	1.708	1.593	E6	9.18
2) SA	Decachlorobiphenyl	5.749	6.580	6.153	5.966	7.087	6.307	E6	8.44
3) L1	AR-1016-1	6.502	6.980	6.039	7.605	7.029	6.831	E4	8.65
4) L1	AR-1016-2	0.923	0.990	0.872	1.106	1.007	0.980	E5	9.07
5) L1	AR-1016-3	5.714	6.120	5.492	6.947	5.832	6.021	E4	9.39
6) L1	AR-1016-4	4.720	5.097	4.409	5.502	5.320	5.009	E4	8.88
7) L1	AR-1016-5	4.827	5.323	4.722	5.788	5.802	5.293	E4	9.67
8) L2	AR-1221-1	1.669					1.669	E4	0.00
9) L2	AR-1221-2	1.189					1.189	E4	0.00
10) L2	AR-1221-3	4.076					4.076	E4	0.00
11) L3	AR-1232-1	4.517					4.517	E4	0.00
12) L3	AR-1232-2	2.511					2.511	E4	0.00
13) L3	AR-1232-3	5.347					5.347	E4	0.00
14) L3	AR-1232-4	2.588					2.588	E4	0.00
15) L3	AR-1232-5	2.040					2.040	E4	0.00
16) L4	AR-1242-1	4.095					4.095	E4	0.00
17) L4	AR-1242-2	5.891					5.891	E4	0.00
18) L4	AR-1242-3	3.726					3.726	E4	0.00
19) L4	AR-1242-4	2.980					2.980	E4	0.00
20) L4	AR-1242-5	4.158					4.158	E4	0.00
21) L5	AR-1248-1	3.160					3.160	E4	0.00
22) L5	AR-1248-2	4.279					4.279	E4	0.00
23) L5	AR-1248-3	5.363					5.363	E4	0.00
24) L5	AR-1248-4	7.213					7.213	E4	0.00
25) L5	AR-1248-5	7.239					7.239	E4	0.00
26) L6	AR-1254-1	7.179					7.179	E4	0.00
27) L6	AR-1254-2	1.230					1.230	E5	0.00
28) L6	AR-1254-3	1.454					1.454	E5	0.00
29) L6	AR-1254-4	1.287					1.287	E5	0.00
30) L6	AR-1254-5	1.519					1.519	E5	0.00
31) L7	AR-1260-1	1.183	1.350	1.197	1.362	1.403	1.299	E5	7.81
32) L7	AR-1260-2	1.505	1.790	1.584	1.680	1.746	1.661	E5	7.02
33) L7	AR-1260-3	1.212	1.475	1.301	1.333	1.542	1.373	E5	9.73
34) L7	AR-1260-4	1.501	1.873	1.640	1.661	1.825	1.700	E5	8.85
35) L7	AR-1260-5	3.464	4.398	3.890	3.745	4.164	3.932	E5	9.23
36) L8	AR-1262-1	2.201					2.201	E5	0.00
37) L8	AR-1262-2	5.112					5.112	E5	0.00
38) L8	AR-1262-3	3.838					3.838	E5	0.00
39) L8	AR-1262-4	1.707					1.707	E5	0.00
40) L8	AR-1262-5	2.910					2.910	E5	0.00
41) L9	AR-1268-1	6.405					6.405	E5	0.00
42) L9	AR-1268-2	6.214					6.214	E5	0.00
43) L9	AR-1268-3	5.301					5.301	E5	0.00
44) L9	AR-1268-4	2.901					2.901	E5	0.00
45) L9	AR-1268-5	2.384					2.384	E6	0.00

Signal #2 Calibration Files

500 =PQ043455.D 1000 =PQ043453.D 750 =PQ043454.D
 250 =PQ043456.D 50 =PQ043457.D

Compound		500	1000	750	250	50	Avg	%RSD	
1) SA	Tetrachloro-m-xylene	5.799	6.039	5.279	6.819	6.350	6.057	E5	9.55
2) SA	Decachlorobiphenyl	2.832	3.255	3.046	2.974	3.554	3.132	E6	8.96
3) L1	AR-1016-1	2.783	2.965	2.583	3.273	3.145	2.950	E4	9.36
4) L1	AR-1016-2	3.719	4.003	3.467	4.291	4.106	3.917	E4	8.32
5) L1	AR-1016-3	1.952	2.177	1.879	2.209	2.301	2.104	E4	8.53

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_Q\Method\
 Method File : PQ091119.M
 Title : GC EXTRACTABLES
 Last Update : Thu Sep 12 02:33:14 2019
 Response Via : Initial Calibration

Calibration Files

500 =PQ043455.D 1000 =PQ043453.D 750 =PQ043454.D
 250 =PQ043456.D 50 =PQ043457.D

	Compound	500	1000	750	250	50	Avg		%RSD
6)	L1 AR-1016-4	1.644	1.807	1.564	1.884	1.812	1.742	E4	7.63
7)	L1 AR-1016-5	2.097	2.296	2.025	2.506	2.266	2.238	E4	8.39
8)	L2 AR-1221-1	5.829					5.829	E3	0.00
9)	L2 AR-1221-2	4.808					4.808	E3	0.00
10)	L2 AR-1221-3	1.641					1.641	E4	0.00
11)	L3 AR-1232-1	1.848					1.848	E4	0.00
12)	L3 AR-1232-2	1.995					1.995	E4	0.00
13)	L3 AR-1232-3	1.085					1.085	E4	0.00
14)	L3 AR-1232-4	1.021					1.021	E4	0.00
15)	L3 AR-1232-5	1.095					1.095	E4	0.00
16)	L4 AR-1242-1	1.728					1.728	E4	0.00
17)	L4 AR-1242-2	2.322					2.322	E4	0.00
18)	L4 AR-1242-3	1.236					1.236	E4	0.00
19)	L4 AR-1242-4	1.271					1.271	E4	0.00
20)	L4 AR-1242-5	2.002					2.002	E4	0.00
21)	L5 AR-1248-1	1.259					1.259	E4	0.00
22)	L5 AR-1248-2	1.771					1.771	E4	0.00
23)	L5 AR-1248-3	1.872					1.872	E4	0.00
24)	L5 AR-1248-4	2.307					2.307	E4	0.00
25)	L5 AR-1248-5	2.834					2.834	E4	0.00
26)	L6 AR-1254-1	4.523					4.523	E4	0.00
27)	L6 AR-1254-2	4.407					4.407	E4	0.00
28)	L6 AR-1254-3	7.459					7.459	E4	0.00
29)	L6 AR-1254-4	5.534					5.534	E4	0.00
30)	L6 AR-1254-5	8.150					8.150	E4	0.00
31)	L7 AR-1260-1	5.251	6.186	5.469	6.139	5.980	5.805	E4	7.24
32)	L7 AR-1260-2	6.785	8.186	7.276	7.837	8.142	7.645	E4	7.88
33)	L7 AR-1260-3	6.120	7.527	6.619	7.436	7.354	7.011	E4	8.77
34)	L7 AR-1260-4	5.806	7.111	6.356	6.627	7.062	6.592	E4	8.19
35)	L7 AR-1260-5	1.646	2.098	1.843	1.830	1.971	1.878	E5	9.00
36)	L8 AR-1262-1	9.887					9.887	E4	0.00
37)	L8 AR-1262-2	2.235					2.235	E5	0.00
38)	L8 AR-1262-3	1.151					1.151	E5	0.00
39)	L8 AR-1262-4	2.253					2.253	E5	0.00
40)	L8 AR-1262-5	1.545					1.545	E5	0.00
41)	L9 AR-1268-1	3.269					3.269	E5	0.00
42)	L9 AR-1268-2	3.199					3.199	E5	0.00
43)	L9 AR-1268-3	2.834					2.834	E5	0.00
44)	L9 AR-1268-4	1.625					1.625	E5	0.00
45)	L9 AR-1268-5	1.256					1.256	E6	0.00

(#) = Out of Range